



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107821

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/18/2020	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/18/2020	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	02/18/2020	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	02/18/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	02/18/2020	17:16
6	L1362-03	1	Solid	20.03	20	20.6	6.90	02/18/2020	17:19
7	L1575-04	1	Solid	20.04	20	21.2	7.69	02/18/2020	17:20
8	L1575-08	1	Solid	20.02	20	21.2	6.56	02/18/2020	17:21
9	L1575-08DUP	1	Solid	20.03	20	21.3	6.57	02/18/2020	17:22
10	CCV2	1	Water	NA	NA	20.1	12.02	02/18/2020	17:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L1573 WorkList ID : 136218 Department : Wet-Chemistry Date : 02-18-2020 15:06:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/18/2020	Solid	L1362-03	Corrosivity	Cool 4 deg C	PSEG05	I63	BU-03-021820	02/18/2020	9045D
02/18/2020	Solid	L1575-04	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-A-S	02/18/2020	9045D
02/18/2020	Solid	L1575-08	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-A-D	02/18/2020	9045D

Date/Time 02-18-2020 16:00
 Received by: Jo
 Relinquished by: Jo

Date/Time 02-18-2020 17:20
 Received by: Jo
 Relinquished by: Jo