


LB 107827

7470A

INSTRUMENT ID: CXL



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	6	0.000	6				
0.2	0.200	0.241	0.041	763	0.0 %	763				
2.5	2.500	2.186	-0.314	8245	0.0 %	8245				
5.0	5.000	5.411	0.411	20653	0.0 %	20653				
7.5	7.500	7.394	-0.106	28284	0.0 %	28284				
10.0	10.000	9.923	-0.077	38011	0.0 %	38011				

LB107827 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Feb 2020 09:03:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	6	-	0.0000	7470A	PPB	19 Feb 2020 09:09:13	S	Std
0.2 - 1	SO.2	763	-	0.2000	7470A	PPB	19 Feb 2020 09:11:20	S	Std
2.5 - 1	SO.5	8245	-	2.5000	7470A	PPB	19 Feb 2020 09:13:27	S	Std
5.0 - 1	SO	20653	-	5.0000	7470A	PPB	19 Feb 2020 09:15:34	S	Std
7.5 - 1	SO	28284	-	7.5000	7470A	PPB	19 Feb 2020 09:17:41	S	Std
10.0 - 1	SO	38011	-	10.0000	7470A	PPB	19 Feb 2020 09:19:47	S	Std
ICV22 - 1	ICV22	15660	4.1133	-	7470A	PPB	19 Feb 2020 09:22:33	U	SMPL
ICB22 - 1	ICB22	-17	0.0387	-	7470A	PPB	19 Feb 2020 09:24:38	U	SMPL
CCV49 - 1	CCV49	19725	5.1698	-	7470A	PPB	19 Feb 2020 09:26:44	U	SMPL
CCB49 - 1	CCB49	-8	0.0410	-	7470A	PPB	19 Feb 2020 09:28:49	U	SMPL
CRA - 1	CRA	789	0.2481	-	7470A	PPB	19 Feb 2020 09:30:55	U	SMPL
HighStd - 1	HighStd	37788	9.8646	-	7470A	PPB	19 Feb 2020 09:33:01	U	SMPL
ChkStd - 1	ChkStd	29236	7.6418	-	7470A	PPB	19 Feb 2020 09:35:07	U	SMPL
PB126906BL - 1	PBW	-34	0.0342	-	7470A	PPB	19 Feb 2020 09:37:14	U	SMPL
PB126906BS - 1	LCSW	16651	4.3708	-	7470A	PPB	19 Feb 2020 09:39:21	U	SMPL
L1558-05 - 1	0109-SITE-WATER	61	0.0589	-	7470A	PPB	19 Feb 2020 09:41:28	U	SMPL
L1558-05DUP - 1	0109-SITE-WATERDUP	126	0.0758	-	7470A	PPB	19 Feb 2020 09:43:35	U	SMPL
L1558-05MS - 1	0109-SITE-WATERMS	17462	4.5816	-	7470A	PPB	19 Feb 2020 09:45:43	U	SMPL
L1558-05MSD - 1	0109-SITE-WATERMSD	17660	4.6331	-	7470A	PPB	19 Feb 2020 09:47:51	U	SMPL
L1558-06 - 1	0110-FIELD-BLANK	-44	0.0316	-	7470A	PPB	19 Feb 2020 09:49:57	U	SMPL
CCV50 - 1	CCV50	19695	5.1620	-	7470A	PPB	19 Feb 2020 09:52:03	U	SMPL
CCB50 - 1	CCB50	-29	0.0355	-	7470A	PPB	19 Feb 2020 09:54:09	U	SMPL
L1566-01 - 1	0119-FIELD-BLANK	11	0.0459	-	7470A	PPB	19 Feb 2020 09:56:15	U	SMPL
L1566-02 - 1	0130-SITE-WATER	119	0.0740	-	7470A	PPB	19 Feb 2020 09:58:21	U	SMPL
L1558-05LX5 - 1		16	0.0472	-	7470A	PPB	19 Feb 2020 10:01:30	U	SMPL
L1558-05A - 1		16118	4.2323	-	7470A	PPB	19 Feb 2020 10:03:36	U	SMPL
CCV51 - 1	CCV51	19813	5.1927	-	7470A	PPB	19 Feb 2020 10:05:42	U	SMPL
CCB51 - 1	CCB51	-13	0.0397	-	7470A	PPB	19 Feb 2020 10:07:49	U	SMPL