



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107828

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/19/2020	10:00
2	CAL2	1	Water	NA	NA	20.1	7.01	02/19/2020	10:01
3	CAL3	1	Water	NA	NA	20.1	10.04	02/19/2020	10:03
4	ICV	1	Water	NA	NA	20.2	7.00	02/19/2020	10:05
5	CCV1	1	Water	NA	NA	20.1	2.01	02/19/2020	10:06
6	L1578-02	1	Water	NA	NA	13.9	6.77	02/19/2020	10:10
7	L1578-02DUP	1	Water	NA	NA	13.8	6.78	02/19/2020	10:11
8	CCV2	1	Water	NA	NA	20.1	12.01	02/19/2020	10:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1578 WorkList ID : 136237 Department : Wet-Chemistry Date : 02-19-2020 09:44:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/17/2020	Water	L1578-02	pH	Cool 4 deg C	PARS15	M31	WFD-WC-L-01	02/17/2020	9040C

Date/Time 02-19-2020 9:50
 Received by: Jo
 Relinquished by: Jo

Date/Time 02-19-2020 10:35
 Received by: Jo
 Relinquished by: Jo