



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107839

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/19/2020	12:50
2	CAL2	1	Water	NA	NA	20.1	7.00	02/19/2020	12:51
3	CAL3	1	Water	NA	NA	20.1	10.03	02/19/2020	12:53
4	ICV	1	Water	NA	NA	20.1	6.99	02/19/2020	13:00
5	CCV1	1	Water	NA	NA	20.2	2.01	02/19/2020	13:02
6	L1581-02	1	Solid	20.03	20	21.5	6.48	02/19/2020	13:15
7	L1581-02DUP	1	Solid	20.04	20	21.6	6.49	02/19/2020	13:16
8	CCV2	1	Water	NA	NA	20.2	12.02	02/19/2020	13:19

13107839

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1581S WorkList ID : 136246 Department : Wet-Chemistry Date : 02-19-2020 12:46:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2020	Solid	L1581-02	pH	Cool 4 deg C	PSEG01	I31	40712	02/19/2020	9045D

Date/Time 02-19-2020 12:45
 Received by: Jo
 Relinquished by: Jo

Date/Time 02-19-2020 14:10
 Received by: Jo
 Relinquished by: Jo