



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107848

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/19/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	02/19/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	02/19/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	02/19/2020	17:05
5	CCV1	1	Water	NA	NA	20.2	2.01	02/19/2020	17:20
6	L1596-01	1	Solid	20.03	20	21.6	5.75	02/19/2020	17:22
7	L1596-01DUP	1	Solid	20.04	20	21.7	5.76	02/19/2020	17:23
8	CCV2	1	Water	NA	NA	20.1	12.02	02/19/2020	17:25

WORKLIST(Hardcopy Internal Chain)

WB107848

WorkList Name : PH L1596

WorkList ID : 136261

Department : Wet-Chemistry

Date : 02-19-2020 15:59:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2020	Solid	L1596-01	pH	Cool 4 deg C	TULL02	M31	FK-PS-01-064	02/19/2020	9045D

Date/Time 02-19-2020 16:00
 Received by: JR
 Relinquished by: JR

Date/Time 02-19-2020 17:10
 Received by: JR
 Relinquished by: JR