

LB107855


7471B

INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

54275



A= 0.0000e+000

B= 1.8292e-004

C= -2.2771e-002

Rho= 0.9998384

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.046	-0.046	-129	0.000	-129				
0.2	0.200	0.178	-0.022	1096	0.0 %	1096				
2.5	2.500	2.562	0.062	14128	0.0 %	14128				
5.0	5.000	4.997	-0.003	27442	0.0 %	27442				
7.5	7.500	7.605	0.105	41699	0.0 %	41699				
10.0	10.000	9.905	-0.095	54275	0.0 %	54275				

LB107855 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 20 Feb 2020 11:24:25					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-129	-	0.0000	7471B	PPB	20 Feb 2020 11:28:12	S	Std
0.2 - 1	50.2	1096	-	0.2000	7471B	PPB	20 Feb 2020 11:30:29	S	Std
2.5 - 1	52.5	14128	-	2.5000	7471B	PPB	20 Feb 2020 11:32:46	S	Std
5.0 - 1	55	27442	-	5.0000	7471B	PPB	20 Feb 2020 11:35:03	S	Std
7.5 - 1	57.5	41699	-	7.5000	7471B	PPB	20 Feb 2020 11:37:18	S	Std
10.0 - 1	510	54275	-	10.0000	7471B	PPB	20 Feb 2020 11:39:34	S	Std
ICV25 - 1	ICV25	22705	4.1305	-7471B	PPB	PPB	20 Feb 2020 11:42:25	U	SMPL
ICB25 - 1	ICB25	-19	-0.0262	-7471B	PPB	PPB	20 Feb 2020 11:44:40	U	SMPL
CCV60 - 1	CCV60	27254	4.9626	-7471B	PPB	PPB	20 Feb 2020 11:46:55	U	SMPL
CCB60 - 1	CCB60	-10	-0.0246	-7471B	PPB	PPB	20 Feb 2020 11:49:09	U	SMPL
CRA - 1	CRA	1101	0.1786	-7471B	PPB	PPB	20 Feb 2020 11:51:24	U	SMPL
HighStd - 1	HighStd	53692	9.7986	-7471B	PPB	PPB	20 Feb 2020 11:53:39	U	SMPL
ChkStd - 1	ChkStd	38361	6.9943	-7471B	PPB	PPB	20 Feb 2020 11:55:54	U	SMPL
PB126927BL - 1	PBS	-4	-0.0235	-7471B	PPB	PPB	20 Feb 2020 11:58:11	U	SMPL
PB126927BS - 1	LCSS	20634	3.7516	-7471B	PPB	PPB	20 Feb 2020 12:00:28	U	SMPL
L1362-01 - 1	BU-03-021820	826	0.1283	-7471B	PPB	PPB	20 Feb 2020 12:02:44	U	SMPL
L1362-01DUP - 1	BU-03-021820DUP	886	0.1393	-7471B	PPB	PPB	20 Feb 2020 12:05:00	U	SMPL
L1362-01MS - 1	BU-03-021820MS	21891	3.9816	-7471B	PPB	PPB	20 Feb 2020 12:07:17	U	SMPL
L1362-01MSD - 1	BU-03-021820MSD	22216	4.0410	-7471B	PPB	PPB	20 Feb 2020 12:09:34	U	SMPL
L1575-03 - 1	TP-A-S	7889	1.4203	-7471B	PPB	PPB	20 Feb 2020 12:11:49	U	SMPL
CCV61 - 1	CCV61	27043	4.9240	-7471B	PPB	PPB	20 Feb 2020 12:14:04	U	SMPL
CCB61 - 1	CCB61	-3	-0.0233	-7471B	PPB	PPB	20 Feb 2020 12:16:20	U	SMPL
L1575-07 - 1	TP-A-D	1574	0.2651	-7471B	PPB	PPB	20 Feb 2020 12:18:35	U	SMPL
L1576-01 - 1	WFD-SB-010203-0-8	7418	1.3341	-7471B	PPB	PPB	20 Feb 2020 12:20:50	U	SMPL
L1576-02 - 1	WFD-SB-1010203-0-8	5682	1.0166	-7471B	PPB	PPB	20 Feb 2020 12:23:05	U	SMPL
L1581-02 - 1	40712	2199	0.3795	-7471B	PPB	PPB	20 Feb 2020 12:25:21	U	SMPL
L1587-01 - 1	0170-COMP-SL-A	9477	1.7108	-7471B	PPB	PPB	20 Feb 2020 12:27:37	U	SMPL
L1587-02 - 1	0171-COMP-SL-B	14442	2.6190	-7471B	PPB	PPB	20 Feb 2020 12:29:52	U	SMPL
L1587-03 - 1	0172-COMP-SL-C	9463	1.7082	-7471B	PPB	PPB	20 Feb 2020 12:32:09	U	SMPL
L1592-01 - 1	FK-BPM-04-005-ABCD	1094	0.1773	-7471B	PPB	PPB	20 Feb 2020 12:34:26	U	SMPL
L1595-01 - 1	FK-BPM-01-C2-C3-C2A-C3A	1275	0.2105	-7471B	PPB	PPB	20 Feb 2020 12:36:43	U	SMPL
L1596-01 - 1	FK-PS-01-064	2760	0.4821	-7471B	PPB	PPB	20 Feb 2020 12:38:59	U	SMPL
CCV62 - 1	CCV62	27099	4.9342	-7471B	PPB	PPB	20 Feb 2020 12:41:14	U	SMPL
CCB62 - 1	CCB62	1	-0.0226	-7471B	PPB	PPB	20 Feb 2020 12:43:30	U	SMPL
L1599-01 - 1	FK-BPM-03-011-ABCD	1329	0.2203	-7471B	PPB	PPB	20 Feb 2020 12:45:45	U	SMPL
L1600-03 - 1	TP-B-S	2011	0.3451	-7471B	PPB	PPB	20 Feb 2020 12:48:00	U	SMPL
L1600-07 - 1	TP-B-D	566	0.0808	-7471B	PPB	PPB	20 Feb 2020 12:50:16	U	SMPL
L1362-01LX5 - 1		190	0.0120	-7471B	PPB	PPB	20 Feb 2020 12:52:32	U	SMPL
L1362-01A - 1		22784	4.1449	-7471B	PPB	PPB	20 Feb 2020 12:54:47	U	SMPL
CCV63 - 1	CCV63	27049	4.9251	-7471B	PPB	PPB	20 Feb 2020 12:59:38	U	SMPL
CCB63 - 1	CCB63	-1	-0.0230	-7471B	PPB	PPB	20 Feb 2020 13:01:53	U	SMPL