



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107858

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/20/2020	09:50
2	L1578-02		1	9	100.00	>212.0	>212.0	02/20/2020	10:20
3	L1578-02DUP		1	9	100.00	>212.0	>212.0	02/20/2020	10:45
4	L1581-01		1	8	100.00	>212.0	>212.0	02/20/2020	11:15
5	L1583-01		1	8	100.00	>212.0	>212.0	02/20/2020	11:42

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

AB107858

WorkList Name : fp22020 WorkList ID : 136291 Department : Wet-Chemistry Date : 02-20-2020 09:21:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2020	Water	L1578-02	Flash Point	Cool 4 deg C	PARS15	M31	WFD-WC-L-01	02/17/2020	1010A
02/26/2020	Water	L1581-01	Flash Point	Cool 4 deg C	PSEG01	I31	40729	02/19/2020	1010A
02/26/2020	Water	L1583-01	Flash Point	Cool 4 deg C	PSEG01	I31	40708	02/19/2020	1010A

Date/Time 02/20/2020 09:25
 Received by: KS
 Relinquished by: SD

Date/Time 02/20/20 11:45
 Received by: JP
 Relinquished by: KS