


LB 107862

ISM02.4

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.1080e-004

C= -1.3787e-002

Rho= 0.9999667

Accept=Accepted

Slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	56	0.000	56					-11
0.05	0.050					252					2
0.20	0.200	0.179	-0.021	916	0.0 %	916					-1
2.50	2.500	2.553	0.053	12176	0.0 %	12176					0
5.00	5.000	4.986	-0.014	23716	0.0 %	23716					-1
7.50	7.500	7.462	-0.038	35462	0.0 %	35462					0
10.0	10.000	10.023	0.023	47612	0.0 %	47612					

LB107862

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	56	-		0 ISM02.4	PPB	2/20/2020 13:50	Std
	0.2 S0.2	916	-		0.2 ISM02.4	PPB	2/20/2020 13:52	Std
	2.5 S2.5	12176	-		2.5 ISM02.4	PPB	2/20/2020 13:54	Std
	5 S5.0	23716	-		5 ISM02.4	PPB	2/20/2020 13:56	Std
	7.5 S7.5	35462	-		7.5 ISM02.4	PPB	2/20/2020 13:59	Std
	10 S10.0	47612	-		10 ISM02.4	PPB	2/20/2020 14:01	Std
ICV	ICV	18711	3.9305 -		ISM02.4	PPB	2/20/2020 14:04	SMPL
ICB	ICB	10	-0.0117 -		ISM02.4	PPB	2/20/2020 14:06	SMPL
CCV064	CCV064	23731	4.9887 -		ISM02.4	PPB	2/20/2020 14:08	SMPL
CCB064	CCB064	10	-0.0117 -		ISM02.4	PPB	2/20/2020 14:10	SMPL
PB126946BL	PBS946	43	-0.0047 -		ISM02.4	PPB	2/20/2020 14:13	SMPL
L1585-01	MCB707	1238	0.2472 -		ISM02.4	PPB	2/20/2020 14:15	SMPL
L1585-02	MCB708	804	0.1557 -		ISM02.4	PPB	2/20/2020 14:17	SMPL
L1585-03	MCB709	681	0.1298 -		ISM02.4	PPB	2/20/2020 14:20	SMPL
L1585-04	MCB710	437	0.0783 -		ISM02.4	PPB	2/20/2020 14:22	SMPL
L1585-05	MCB711	379	0.0661 -		ISM02.4	PPB	2/20/2020 14:24	SMPL
L1585-06	MCB712	325	0.0547 -		ISM02.4	PPB	2/20/2020 14:26	SMPL
L1585-07	MCB713	271	0.0433 -		ISM02.4	PPB	2/20/2020 14:29	SMPL
L1585-08	MCB714	865	0.1686 -		ISM02.4	PPB	2/20/2020 14:31	SMPL
L1585-09	MCB715	510	0.0937 -		ISM02.4	PPB	2/20/2020 14:33	SMPL
L1585-10	MCB716	1265	0.2529 -		ISM02.4	PPB	2/20/2020 14:35	SMPL
L1585-11	MCB717	635	0.1201 -		ISM02.4	PPB	2/20/2020 14:38	SMPL
L1585-12	MCB718	1582	0.3197 -		ISM02.4	PPB	2/20/2020 14:40	SMPL
L1585-13	MCB718D	1544	0.3117 -		ISM02.4	PPB	2/20/2020 14:42	SMPL
L1585-14	MCB718S	13505	2.833 -		ISM02.4	PPB	2/20/2020 14:44	SMPL
CCV065	CCV065	23763	4.9954 -		ISM02.4	PPB	2/20/2020 14:47	SMPL
CCB065	CCB065	21	-0.0094 -		ISM02.4	PPB	2/20/2020 14:49	SMPL