



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107866

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/20/2020	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	02/20/2020	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	02/20/2020	15:02
4	ICV	1	Water	NA	NA	20.1	6.99	02/20/2020	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	02/20/2020	15:06
6	L1603-02	1	Water	NA	NA	17.7	6.49	02/20/2020	15:09
7	L1603-02DUP	1	Water	NA	NA	17.8	6.50	02/20/2020	15:10
8	L1610-01	1	Water	NA	NA	13.3	7.12	02/20/2020	15:11
9	CCV2	1	Water	NA	NA	20.1	12.02	02/20/2020	15:12

2B107866

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1610 WorkList ID : 136300 Department : Wet-Chemistry Date : 02-20-2020 13:32:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2020	Water	L1603-02	pH	Cool 4 deg C	PARS15	M41	HAM-WC-L-02	02/19/2020	9040C
02/20/2020	Water	L1610-01	pH	Cool 4 deg C	RAIL01	M31	T1-SECTION-C	02/20/2020	9040C

Date/Time 02-20-2020 13:35
Received by: JP
Relinquished by: JP

Date/Time 02-20-2020 16:30
Received by: JP
Relinquished by: JP