



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 02/21/2020OVEN IN TEMP1: 98 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 02/20/2020 17:00 DATE & TIME OUT1: 02/21/2020 07:20OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 02/21/2020 08:00 DATE & TIME OUT2: 02/21/2020 09:35Run Number: LB107867BalanceID: WC SC-3OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB107867BL	LB107867BL	85.2147	85.2147	100	85.2147	85.2147	85.2147	0.0000	0
2	L1610-01	T1-SECTION-C	90.1565	90.1565	100	90.7242	90.7242	90.7242	0.5677	5677
3	L1610-01DUP	T1-SECTION-CDUP	90.0459	90.0459	100	90.6130	90.6130	90.6130	0.5671	5671

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

18107867

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS L1610 WATER WorkList ID : 136304 Department : Wet-Chemistry Date : 02-20-2020 13:37:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2020	Water	L1610-01-2	TS	Cool 4 deg C	RAIL01	M31	T1-SECTION-C	02/20/2020	SM2540 B

Date/Time 02.20.2020 16:00
Received by: 
Relinquished by:

Date/Time 02.20.2020 16:00
Received by: 
Relinquished by: