



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107872

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/20/2020	12:20
2	L1583-02		1	8	55.00	131.0	131.0	02/20/2020	12:50
3	L1583-02DUP		1	8	56.00	132.8	132.8	02/20/2020	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 107872

WorkList Name : fp22020-1 WorkList ID : 136292 Department : Wet-Chemistry Date : 02-20-2020 09:22:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/26/2020	Solid	L1583-02 A	Flash Point	Cool 4 deg C	PSEG01	I31	13192	02/19/2020	1010A

Date/Time 02/20/20 11:45
 Received by: KS
 Relinquished by: SP

Date/Time 02/20/20 13:25
 Received by: SP
 Relinquished by: KS