



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107877

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/20/2020	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	02/20/2020	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	02/20/2020	15:02
4	ICV	1	Water	NA	NA	20.1	6.99	02/20/2020	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	02/20/2020	15:06
6	L1618-03	1	Water	NA	NA	14.8	9.62	02/20/2020	15:30
7	L1618-03DUP	1	Water	NA	NA	14.9	9.63	02/20/2020	15:31
8	CCV2	1	Water	NA	NA	20.1	12.02	02/20/2020	15:33

VB 107877
11

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1631 WorkList ID : 136311 Department : Wet-Chemistry Date : 02-20-2020 15:53:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/20/2020	Water	L1618-03	Corrosivity	Cool 4 deg C	DVIR01	M31	TANK-MOAT-DISCHARGE-(E	02/20/2020	9040C

Date/Time 02.20.2020 16:00
Received by: JS
Relinquished by: JS

Date/Time 02.20.2020 17:30
Received by: KS
Relinquished by: Jn