



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107878

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/20/2020	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	02/20/2020	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	02/20/2020	15:02
4	ICV	1	Water	NA	NA	20.1	6.99	02/20/2020	15:05
5	CCV1	1	Water	NA	NA	20.2	2.01	02/20/2020	17:00
6	L1603-01	1	Solid	20.03	20	20.9	8.54	02/20/2020	17:02
7	L1603-01DUP	1	Solid	20.04	20	20.8	8.55	02/20/2020	17:03
8	L1631-04	1	Solid	20.03	20	20.5	8.28	02/20/2020	17:05
9	L1631-08	1	Solid	20.05	20	20.7	8.05	02/20/2020	17:06
10	L1631-12	1	Solid	20.03	20	20.6	7.51	02/20/2020	17:10
11	CCV2	1	Water	NA	NA	20.2	12.02	02/20/2020	17:11

WORKLIST(Hardcopy Internal Chain)

107878

WorkList Name: CORROSIONITY L1618 WorkList ID: 136312 Department: Wet-Chemistry Date: 02-20-2020 15:54:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2020	Solid	L1603-01	Corrosivity	Cool 4 deg C	PARS15	M41	HAM-WC-S-SB-010203	02/19/2020	9045D
02/20/2020	Solid	L1631-04	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-C-S	02/20/2020	9045D
02/20/2020	Solid	L1631-08	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-D-S	02/20/2020	9045D
02/20/2020	Solid	L1631-12	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-D-D	02/20/2020	9045D

Date/Time 02-20-2020 16:00
 Received by: J9
 Relinquished by: KS

Date/Time 02-20-2020 17:40
 Received by: KS
 Relinquished by: SN