



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107890

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/21/2020	08:55
2	L1603-02		1	9	100.00	>212.0	>212.0	02/21/2020	09:20
3	L1603-02DUP		1	9	100.00	>212.0	>212.0	02/21/2020	09:45
4	L1610-01		1	8	100.00	>212.0	>212.0	02/21/2020	10:10
5	L1618-03		1	9	100.00	>212.0	>212.0	02/21/2020	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB107890

WorkList Name : FP22120 WorkList ID : 136324 Department : Wet-Chemistry Date : 02-20-2020 17:18:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2020	Water	L1603-02	Flash Point	Cool 4 deg C	PARS15	M41	HAM-WC-L-02	02/19/2020	1010A
03/05/2020	Water	L1610-01	Flash Point	Cool 4 deg C	RAIL01	M31	T1-SECTION-C	02/20/2020	1010A
02/21/2020	Water	L1618-03	Flash Point	Cool 4 deg C	DVIR01	M31	TANK-MOAT-DISCHARGE-(E	02/20/2020	1010A

Date/Time 02/21/20 08:30
 Received by: KS
 Relinquished by: KS

Date/Time 02/21/20 10:45
 Received by: KS
 Relinquished by: KS