


LB107891

ISM02.4

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.2498e-004

C= -5.1228e-002

Rho= 0.9997604

Accept=Accepted

*Slope
Intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>σ/0.0</i>
0.00	0.000	-0.015	-0.015	160	0.000	160					-
0.05	0.050					252					-12
0.20	0.200	0.177	-0.023	1015	0.0 %	1015					1
2.50	2.500	2.524	0.024	11446	0.0 %	11446					2
5.00	5.000	5.121	0.121	22989	0.0 %	22989					-2
7.50	7.500	7.353	-0.147	32911	0.0 %	32911					0
10.0	10.000	10.040	0.040	44856	0.0 %	44856					

LB107891

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	160	-		0 ISM02.4	PPB	2/21/2020 11:37	Std
	0.2 S0.2	1015	-		0.2 ISM02.4	PPB	2/21/2020 11:39	Std
	2.5 S2.5	11446	-		2.5 ISM02.4	PPB	2/21/2020 11:41	Std
	5 S5.0	22989	-		5 ISM02.4	PPB	2/21/2020 11:43	Std
	7.5 S7.5	32911	-		7.5 ISM02.4	PPB	2/21/2020 11:46	Std
	10 S10.0	44856	-		10 ISM02.4	PPB	2/21/2020 11:48	Std
ICV	ICV	17274	3.835 -		ISM02.4	PPB	2/21/2020 11:51	SMPL
ICB	ICB	138	-0.0202 -		ISM02.4	PPB	2/21/2020 11:53	SMPL
CCV070	CCV070	22713	5.0587 -		ISM02.4	PPB	2/21/2020 11:55	SMPL
CCB070	CCB070	123	-0.0236 -		ISM02.4	PPB	2/21/2020 11:58	SMPL
PB126986BL	PBW986	279	0.0115 -		ISM02.4	PPB	2/21/2020 12:00	SMPL
L1609-07	MX2L77	4743	1.0158 -		ISM02.4	PPB	2/21/2020 12:02	SMPL
L1609-08	MX2L77D	4561	0.9749 -		ISM02.4	PPB	2/21/2020 12:04	SMPL
L1609-09	MX2L77S	9102	1.9965 -		ISM02.4	PPB	2/21/2020 12:07	SMPL
CCV071	CCV071	22571	5.0268 -		ISM02.4	PPB	2/21/2020 12:11	SMPL
CCB071	CCB071	144	-0.0188 -		ISM02.4	PPB	2/21/2020 12:13	SMPL