


LB107897

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	755	0.000	755				
0.2	0.200	0.196	-0.004	1777	0.0 %	1777				
2.5	2.500	2.485	-0.015	14474	0.0 %	14474				
5.0	5.000	4.997	-0.003	28404	0.0 %	28404				
7.5	7.500	7.521	0.021	42400	0.0 %	42400				
10.0	10.000	9.989	-0.011	56088	0.0 %	56088				

LB107897 INSTRUMENT ID : CV1

Method: 7471B	Operator: Admin	Date of Analysis: 21 Feb 2020 12:24:53							
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	755	-	0.0000	7471B	PPB	21 Feb 2020 12:26:38	S	Std
0.2 - 1	50.2	1777	-	0.2000	7471B	PPB	21 Feb 2020 12:28:55	S	Std
2.5 - 1	52.5	14474	-	2.5000	7471B	PPB	21 Feb 2020 12:31:12	S	Std
5.0 - 1	55	28404	-	5.0000	7471B	PPB	21 Feb 2020 12:33:29	S	Std
7.5 - 1	57.5	42400	-	7.5000	7471B	PPB	21 Feb 2020 12:35:45	S	Std
10.0 - 1	58	56088	-	10.0000	7471B	PPB	21 Feb 2020 12:38:01	S	Std
ICV27 - 1	ICV27	21863	3.8177	-	7471B	PPB	21 Feb 2020 12:40:40	U	SMPL
ICB27 - 1	ICB27	736	0.0080	-	7471B	PPB	21 Feb 2020 12:42:55	U	SMPL
CCV72 - 1	CCV72	28545	5.0226	-	7471B	PPB	21 Feb 2020 12:45:10	U	SMPL
CCB72 - 1	CCB72	723	0.0057	-	7471B	PPB	21 Feb 2020 12:47:24	U	SMPL
CRA - 1	CRA	1768	0.1941	-	7471B	PPB	21 Feb 2020 12:49:40	U	SMPL
HighStd - 1	HighStd	56260	10.0203	-	7471B	PPB	21 Feb 2020 12:51:54	U	SMPL
ChkStd - 1	ChkStd	38026	6.7323	-	7471B	PPB	21 Feb 2020 12:54:09	U	SMPL
PB126973BL - 1	PBS	695	0.0006	-	7471B	PPB	21 Feb 2020 12:56:26	U	SMPL
PB126973BS - 1	LCSS	21096	3.6794	-	7471B	PPB	21 Feb 2020 12:58:42	U	SMPL
L1526-03 - 1	E156-B4-021820-COMP-0-5	42080	7.4633	-	7471B	PPB	21 Feb 2020 13:00:58	U	SMPL
L1526-05 - 1	E156-B4-021820-COMP-5-10	389	-0.0546	-	7471B	PPB	21 Feb 2020 13:03:15	U	SMPL
L1526-07 - 1	E156-B4-021820-COMP-10-15	142	-0.0991	-	7471B	PPB	21 Feb 2020 13:05:32	U	SMPL
L1526-09 - 1	E156-B4-021820-COMP-BD	85	-0.1094	-	7471B	PPB	21 Feb 2020 13:07:49	U	SMPL
L1604-06 - 1	HAM-SB-010203-5-52	924	0.0419	-	7471B	PPB	21 Feb 2020 13:10:04	U	SMPL
CCV73 - 1	CCV73	28488	5.0124	-	7471B	PPB	21 Feb 2020 13:12:19	U	SMPL
CCB73 - 1	CCB73	718	0.0048	-	7471B	PPB	21 Feb 2020 13:14:34	U	SMPL
L1606-02 - 1	361	1963	0.2293	-	7471B	PPB	21 Feb 2020 13:16:49	U	SMPL
L1606-02DUP - 1	361DUP	2058	0.2464	-	7471B	PPB	21 Feb 2020 13:19:05	U	SMPL
L1606-02MS - 1	361MS	22861	3.9977	-	7471B	PPB	21 Feb 2020 13:21:20	U	SMPL
L1606-02MSD - 1	361MSD	23923	4.1892	-	7471B	PPB	21 Feb 2020 13:23:35	U	SMPL
L1613-01 - 1	TP-14-EP1-2	58492	10.4228	-	7471B	PPB	21 Feb 2020 13:25:50	U	SMPL
L1613-02 - 1	TP-14-EP2-3	2329	0.2953	-	7471B	PPB	21 Feb 2020 13:28:06	U	SMPL
L1625-01 - 1	FK-GF-02-050-A	9225	1.5388	-	7471B	PPB	21 Feb 2020 13:30:24	U	SMPL
L1625-03 - 1	FK-GF-02-050-B	7441	1.2171	-	7471B	PPB	21 Feb 2020 13:32:40	U	SMPL
L1626-01 - 1	FK-GF-02-051-A	6955	1.1294	-	7471B	PPB	21 Feb 2020 13:34:57	U	SMPL
L1626-03 - 1	FK-GF-02-051-B	7703	1.2643	-	7471B	PPB	21 Feb 2020 13:37:12	U	SMPL
CCV74 - 1	CCV74	28502	5.0149	-	7471B	PPB	21 Feb 2020 13:39:27	U	SMPL
CCB74 - 1	CCB74	738	0.0084	-	7471B	PPB	21 Feb 2020 13:41:43	U	SMPL
L1631-03 - 1	TP-C-S	25010	4.3852	-	7471B	PPB	21 Feb 2020 13:43:58	U	SMPL
L1631-07 - 1	TP-D-S	3122	0.4383	-	7471B	PPB	21 Feb 2020 13:46:13	U	SMPL
L1631-11 - 1	TP-D-D	6735	1.0898	-	7471B	PPB	21 Feb 2020 13:48:28	U	SMPL
L1606-02LX5 - 1		1000	0.0556	-	7471B	PPB	21 Feb 2020 13:50:44	U	SMPL
L1606-02A - 1		24079	4.2173	-	7471B	PPB	21 Feb 2020 13:53:00	U	SMPL
L1613-01DLX2 - 1		29881	5.2635	-	7471B	PPB	21 Feb 2020 13:55:15	U	SMPL
CCV75 - 1	CCV75	28573	5.0277	-	7471B	PPB	21 Feb 2020 13:57:32	U	SMPL
CCB75 - 1	CCB75	730	0.0069	-	7471B	PPB	21 Feb 2020 13:59:49	U	SMPL