


LB107908

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	120	0.000	120				
0.2	0.200	0.215	0.015	945	0.0 %	945				
2.5	2.500	2.538	0.038	10020	0.0 %	10020				
5.0	5.000	4.958	-0.042	19471	0.0 %	19471				
7.5	7.500	7.400	-0.100	29013	0.0 %	29013				
10.0	10.000	10.084	0.084	39497	0.0 %	39497				

LB107908 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 21 Feb 2020 14:52:05					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	120	-	0.0000	7470A	PPB	21 Feb 2020 14:55:25	S	Std
0.2 - 1	50.2	945	-	0.2000	7470A	PPB	21 Feb 2020 14:57:33	S	Std
2.5 - 1	52.5	10020	-	2.5000	7470A	PPB	21 Feb 2020 14:59:41	S	Std
5.0 - 1	55	19471	-	5.0000	7470A	PPB	21 Feb 2020 15:01:48	S	Std
7.5 - 1	57.5	29013	-	7.5000	7470A	PPB	21 Feb 2020 15:03:55	S	Std
10.0 - 1	510	39497	-	10.0000	7470A	PPB	21 Feb 2020 15:06:01	S	Std
ICV28 - 1	ICV28	14307	3.6358	-	7470A	PPB	21 Feb 2020 15:08:35	U	SMPL
ICB28 - 1	ICB28	73	-0.0078	-	7470A	PPB	21 Feb 2020 15:10:40	U	SMPL
CCV76 - 1	CCV76	19340	4.9242	-	7470A	PPB	21 Feb 2020 15:12:46	U	SMPL
CCB76 - 1	CCB76	78	-0.0065	-	7470A	PPB	21 Feb 2020 15:14:52	U	SMPL
CRA - 1	CRA	942	0.2146	-	7470A	PPB	21 Feb 2020 15:16:57	U	SMPL
HighStd - 1	HighStd	39985	10.2089	-	7470A	PPB	21 Feb 2020 15:19:03	U	SMPL
ChkStd - 1	ChkStd	26711	6.8110	-	7470A	PPB	21 Feb 2020 15:21:09	U	SMPL
PB127005BL - 1	PBW	44	-0.0152	-	7470A	PPB	21 Feb 2020 15:23:15	U	SMPL
PB127005BS - 1	LCSW	14223	3.6143	-	7470A	PPB	21 Feb 2020 15:25:21	U	SMPL
L1603-02 - 1	HAM-WC-L-02	4017	1.0018	-	7470A	PPB	21 Feb 2020 15:27:28	U	SMPL
L1603-02DUP - 1	HAM-WC-L-02DUP	3939	0.9818	-	7470A	PPB	21 Feb 2020 15:29:35	U	SMPL
L1603-02MS - 1	HAM-WC-L-02MS	18501	4.7094	-	7470A	PPB	21 Feb 2020 15:31:42	U	SMPL
L1603-02MSD - 1	HAM-WC-L-02MSD	19057	4.8517	-	7470A	PPB	21 Feb 2020 15:33:50	U	SMPL
L1605-01 - 1	HAM-GW-02-10	7203	1.8173	-	7470A	PPB	21 Feb 2020 15:35:56	U	SMPL
CCV77 - 1	CCV77	19647	5.0028	-	7470A	PPB	21 Feb 2020 15:38:01	U	SMPL
CCB77 - 1	CCB77	81	-0.0058	-	7470A	PPB	21 Feb 2020 15:40:07	U	SMPL
L1610-01 - 1	T1-SECTION-C	16932	4.3078	-	7470A	PPB	21 Feb 2020 15:42:13	U	SMPL
L1603-02LX5 - 1		792	0.1762	-	7470A	PPB	21 Feb 2020 15:44:19	U	SMPL
L1603-02A - 1		21057	5.3637	-	7470A	PPB	21 Feb 2020 15:46:25	U	SMPL
CCV78 - 1	CCV78	19823	5.0478	-	7470A	PPB	21 Feb 2020 15:48:32	U	SMPL
CCB78 - 1	CCB78	93	-0.0027	-	7470A	PPB	21 Feb 2020 15:50:38	U	SMPL