


LB107911

245.1

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.7189e-004

C= -8.3202e-003

Rho= 0.9999993

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	40	0.000	40				
0.05	0.050	0.046	-0.004	198	0.0 %	198				
0.2	0.200	0.197	-0.003	755	0.0 %	755				
2.5	2.500	2.508	0.008	9256	0.0 %	9256				
5.0	5.000	4.996	-0.004	18405	0.0 %	18405				
7.5	7.500	7.503	0.003	27625	0.0 %	27625				
10.0	10.000	9.998	-0.002	36802	0.0 %	36802				

LB107911 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Feb 2020 16:01:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	40	-	0.0000	245.1	PPB	21 Feb 2020 16:06:00	S	Std
0.05 - 1	SO.05	198	-	0.0500	245.1	PPB	21 Feb 2020 16:08:17	S	Std
0.2 - 1	SO.2	755	-	0.2000	245.1	PPB	21 Feb 2020 16:10:34	S	Std
2.5 - 1	SO.25	9256	-	2.5000	245.1	PPB	21 Feb 2020 16:12:51	S	Std
5.0 - 1	SO.5	18405	-	5.0000	245.1	PPB	21 Feb 2020 16:15:07	S	Std
7.5 - 1	SO.75	27625	-	7.5000	245.1	PPB	21 Feb 2020 16:17:23	S	Std
10.0 - 1	SO.10	36802	-	10.0000	245.1	PPB	21 Feb 2020 16:19:38	S	Std
ICV29 - 1	ICV29	14377	3.9007	-245.1		PPB	21 Feb 2020 16:22:24	U	SMPL
ICB29 - 1	ICB29	17	-0.0037	-245.1		PPB	21 Feb 2020 16:24:39	U	SMPL
CCV79 - 1	CCV79	18786	5.0995	-245.1		PPB	21 Feb 2020 16:26:54	U	SMPL
CCB79 - 1	CCB79	6	-0.0067	-245.1		PPB	21 Feb 2020 16:29:08	U	SMPL
CRA - 1	CRA	190	0.0433	-245.1		PPB	21 Feb 2020 16:31:23	U	SMPL
HighStd - 1	HighStd	36829	10.0053	-245.1		PPB	21 Feb 2020 16:33:38	U	SMPL
ChkStd - 1	ChkStd	27859	7.5664	-245.1		PPB	21 Feb 2020 16:35:53	U	SMPL
PB127019BL - 1	PBW	-33	-0.0173	-245.1		PPB	21 Feb 2020 16:38:09	U	SMPL
PB127019BS - 1	LCSW	15289	4.1487	-245.1		PPB	21 Feb 2020 16:40:24	U	SMPL
L1643-02 - 1	EFFLUENT-METALS	176	0.0395	-245.1		PPB	21 Feb 2020 16:47:33	U	SMPL
L1643-02DUP - 1	EFFLUENT-METALSDUP	156	0.0341	-245.1		PPB	21 Feb 2020 16:49:49	U	SMPL
L1643-02MS - 1	EFFLUENT-METALSMS	12482	3.3855	-245.1		PPB	21 Feb 2020 17:03:06	U	SMPL
L1643-02MSD - 1	EFFLUENT-METALSMSD	12689	3.4418	-245.1		PPB	21 Feb 2020 17:05:23	U	SMPL
CCV80 - 1	CCV80	19426	5.2735	-245.1		PPB	21 Feb 2020 17:10:14	U	SMPL
CCB80 - 1	CCB80	-343	-0.1016	-245.1		PPB	21 Feb 2020 17:12:29	U	SMPL
L1643-02LX5 - 1		-11	-0.0113	-245.1		PPB	21 Feb 2020 17:17:32	U	SMPL
L1643-02A - 1		12258	3.3246	-245.1		PPB	21 Feb 2020 17:19:47	U	SMPL
CCV81 - 1	CCV81	17019	4.6191	-245.1		PPB	21 Feb 2020 17:22:02	U	SMPL
CCB81 - 1	CCB81	-291	-0.0874	-245.1		PPB	21 Feb 2020 17:24:18	U	SMPL