



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107912

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	02/21/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	02/21/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/21/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.01	02/21/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	02/21/2020	17:16
6	L1657-04	1	Solid	20.03	20	21.7	8.59	02/21/2020	17:19
7	L1657-08	1	Solid	20.04	20	21.6	7.27	02/21/2020	17:20
8	L1659-01	1	Solid	20.02	20	19.8	7.04	02/21/2020	17:22
9	L1659-04	1	Solid	20.03	20	20.4	7.05	02/21/2020	17:25
10	L1659-04DUP	1	Solid	20.04	20	20.5	7.06	02/21/2020	17:26
11	CCV2	1	Water	NA	NA	20.1	12.02	02/21/2020	17:30

WORKLIST(Hardcopy Internal Chain)

LB 107912

WorkList Name : CORROSIONITY L1604 WorkList ID : 136390 Department : Wet-Chemistry Date : 02-21-2020 16:49:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/21/2020	Solid	L1657-04	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-F-S	02/21/2020	9045D
02/21/2020	Solid	L1657-08	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-F-D	02/21/2020	9045D
02/21/2020	Solid	L1659-01	Corrosivity	Cool 4 deg C	PSEG01	I32	WH-1-COMP-1	02/21/2020	9045D
02/21/2020	Solid	L1659-04	Corrosivity	Cool 4 deg C	PSEG01	I32	WH-2-COMP-2	02/21/2020	9045D

Date/Time 02.21.2020/11:50
 Received by: JA
 Relinquished by: JA

Date/Time 02.21.2020 18:00
 Received by: DA
 Relinquished by: JA