



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107913

Slope : 97.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	02/21/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	02/21/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/21/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.01	02/21/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.02	02/21/2020	17:33
6	L1593-01	1	Solid	20.03	20	19.6	5.16	02/21/2020	17:35
7	L1593-01DUP	1	Solid	20.02	20	19.7	5.17	02/21/2020	17:36
8	CCV2	1	Water	NA	NA	20.1	12.02	02/21/2020	17:38

WORKLIST(Hardcopy Internal Chain)

2B10f913

WorkList Name : PH L1693 WorkList ID : 136389 Department : Wet-Chemistry Date : 02-21-2020 16:49:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/19/2020	Solid	L1593-01	pH	Cool 4 deg C	TULL01	M31	STOCKPILE	02/19/2020	9045D

Date/Time 02-21-2020 16:50
 Received by: JP
 Relinquished by: DK

Date/Time 02-21-2020 17:40
 Received by: DK
 Relinquished by: JP