



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/25/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 02/24/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 02/25/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107939

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1367-05	OK-02-022420	1	1.12	9.71	8.93	90.9	
L1367-06	OK-02-022420	2	1.17	9.77	8.8	88.7	
L1367-07	OK-02-022420-1	3	1.19	9.91	8.6	85.0	
L1660-01	01-A-01-B-01-C	4	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1660-02	02-A-02-B-02-C	5	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1660-03	03-A-03-B-03-C	6	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1665-01	S305-T306-K297-CLEAN	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1665-02	S305-STAIN	8	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1665-03	T306-STAIN	9	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1665-04	K297-STAIN	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1669-01	VNJ-231	11	1.16	9.51	8.47	87.5	
L1669-03	72-11991	12	1.18	9.69	8.75	89.0	
L1669-05	RBR-2520699	13	1.18	9.94	8.43	82.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$