



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107940

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/24/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	02/24/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	02/24/2020	17:03
4	ICV	1	Water	NA	NA	20.2	7.00	02/24/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	02/24/2020	17:07
6	L1367-08	1	Solid	20.03	20	20.9	8.70	02/24/2020	17:09
7	L1367-08DUP	1	Solid	20.04	20	20.8	8.71	02/24/2020	17:11
8	CCV2	1	Water	NA	NA	20.1	12.02	02/24/2020	17:12

WORKLIST(Hardcopy Internal Chain)

185107940

WorkList Name : CORROSIVITY L1369 WorkList ID : 136430 Department : Wet-Chemistry Date : 02-24-2020 15:32:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/24/2020	Solid	L1367-08	Corrosivity	Cool 4 deg C	PSEG05		OK-02-022420	02/24/2020	9045D

Date/Time 02-24-2020 16110
 Received by: SD
 Relinquished by: SD

Date/Time 02-24-2020 17110
 Received by: SD
 Relinquished by: SD