



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107941

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/24/2020	16:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/24/2020	16:11
3	CAL3	1	Water	NA	NA	20.2	10.03	02/24/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.01	02/24/2020	16:15
5	CCV1	1	Water	NA	NA	20.2	2.01	02/24/2020	16:16
6	L1666-01	1	Water	NA	NA	18.1	7.85	02/24/2020	16:20
7	L1666-02	1	Water	NA	NA	18.2	5.16	02/24/2020	16:22
8	L1668-03	1	Water	NA	NA	19.0	7.03	02/24/2020	16:24
9	L1668-09	1	Water	NA	NA	18.9	6.32	02/24/2020	16:25
10	L1670-09	1	Water	NA	NA	19.9	7.23	02/24/2020	16:26
11	L1670-10	1	Water	NA	NA	19.7	7.33	02/24/2020	16:30
12	L1670-10DUP	1	Water	NA	NA	19.8	7.34	02/24/2020	16:31
13	CCV2	1	Water	NA	NA	20.1	12.02	02/24/2020	16:33

VB 107941

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1370 WATER WorkList ID : 136431 Department : Wet-Chemistry Date : 02-24-2020 15:33:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/24/2020	Water	L1666-01	pH	Cool 4 deg C	PSEG01	I33	NB-21944	02/24/2020	9040C
02/24/2020	Water	L1666-02	pH	Cool 4 deg C	PSEG01	I33	NB-30871	02/24/2020	9040C
02/24/2020	Water	L1668-03	pH	Cool 4 deg C	PSEG01	I33	30893	02/24/2020	9040C
02/24/2020	Water	L1668-09	pH	Cool 4 deg C	PSEG01	I33	30873	02/24/2020	9040C
02/24/2020	Water	L1670-09	pH	Cool 4 deg C	PSEG01	I33	30892	02/24/2020	9040C
02/24/2020	Water	L1670-10	pH	Cool 4 deg C	PSEG01	I33	30884	02/24/2020	9040C

Date/Time 02-24-2020 15:40
Received by: JS
Relinquished by: JS

Date/Time 02-24-2020 18:00
Received by: JS
Relinquished by: JS