



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107946

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	02/25/2020	09:20
2	L1666-01		1	9	100.00	>212.0	>212.0	02/25/2020	09:45
3	L1666-01DUP		1	9	100.00	>212.0	>212.0	02/25/2020	10:10
4	L1666-02		1	8	100.00	>212.0	>212.0	02/25/2020	10:35
5	L1668-03		1	9	100.00	>212.0	>212.0	02/25/2020	10:58
6	L1668-09		1	8	100.00	>212.0	>212.0	02/25/2020	11:25
7	L1670-13		1	9	100.00	>212.0	>212.0	02/25/2020	11:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB107946

WorkList Name : fp22420 WorkList ID : 136411 Department : Wet-Chemistry Date : 02-24-2020 09:08:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2020	Water	L1666-01	Flash Point	Cool 4 deg C	PSEG01	I33	NB-21944	02/24/2020	1010A
03/09/2020	Water	L1666-02	Flash Point	Cool 4 deg C	PSEG01	I33	NB-30871	02/24/2020	1010A
03/09/2020	Water	L1668-03	Flash Point	Cool 4 deg C	PSEG01	I33	30893	02/24/2020	1010A
03/09/2020	Water	L1668-09	Flash Point	Cool 4 deg C	PSEG01	I33	30873	02/24/2020	1010A
03/09/2020	Water	L1670-13	Flash Point	Cool 4 deg C	PSEG01	I33	WATER-COMP	02/24/2020	1010A

Date/Time 02/25/20 08:45
 Received by: KS
 Relinquished by: JS

Date/Time 02/25/20 12:00
 Received by: JS
 Relinquished by: KS