



PERCENT SOLID

Supervisor: apatel
Analyst: Daniel
Date: 2/26/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 02/25/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 02/26/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107951

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1370-06	OR-03-022520	22	1.18	9.69	9.18	94.0	
L1370-07	OR-03-022520	23	1.17	9.86	9.27	93.2	
L1370-08	OR-03-022520-1	24	1.17	9.79	9.28	94.1	
L1370-09	OR-03-022520-2	25	1.13	9.62	9.32	96.5	
L1573-08	LOC-1	14	1.18	9.81	8.96	90.2	
L1573-09	LOC-2	15	1.14	9.54	7.79	79.2	
L1573-10	LOC-3	16	1.15	9.77	7.86	77.8	
L1668-01	30875	1	1.00	2.00	2.00	100.0	DEBIS 100% SOLIDS
L1668-05	30830	2	1.15	9.96	7.77	75.1	
L1668-07	30793	3	1.00	2.00	2.00	100.0	DEBIS 100% SOLIDS
L1670-11	SOLID-COMP	4	1.00	2.00	2.00	100.0	DEBIS 100% SOLIDS
L1674-02	P002-SS001-0006-01	5	1.17	9.63	7.89	79.4	
L1674-03	L1674-02MS	6	1.17	9.63	7.89	79.4	
L1674-04	L1674-02MSD	7	1.17	9.63	7.89	79.4	
L1674-05	P002-SS001-0612-01	8	1.17	9.83	8.8	88.1	
L1674-06	P002-SS002-0006-01	9	1.13	9.97	8.46	82.9	
L1674-07	P002-SS002-0612-01	10	1.17	9.86	8.76	87.3	
L1674-08	P002-SS003-0006-01	11	1.17	9.85	8.25	81.6	
L1674-09	P002-SS003-0612-01	12	1.14	9.79	8.59	86.1	
L1674-10	P002-SS003-0612-02	13	1.18	9.97	8.78	86.5	
L1675-01	COMP-FRONT-PARKING	29	1.12	9.56	8.48	87.2	
L1675-02	GRAB-FRONT-PARKING-LOT	30	1.14	9.93	8.96	89.0	
L1676-01	DRUMS-1-2	26	1.00	2.00	2.00	100.0	OILY DEBRIS, 100% SOLIDS
L1676-03	DRUMS-3-4	27	1.00	2.00	2.00	100.0	OILY DEBRIS, 100% SOLIDS
L1676-05	RO-1-2	28	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1678-01	VNJ-241	17	1.13	9.91	8.99	89.5	
L1678-03	VNJ-227	18	1.00	2.00	2.00	100.0	100% SOLIDS
L1679-01	SB-1-12-12.5	31	1.19	9.75	8.54	85.9	



PERCENT SOLID

Supervisor: apatel
Analyst: Daniel
Date: 2/26/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 02/25/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 02/26/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107951

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1679-02	SB-2-12.5-13	32	1.17	9.53	8.43	86.8	
L1679-03	SB-3-12-12.5	33	1.18	9.74	8.71	88.0	
L1679-04	SB-4-12.5-13	34	1.1	9.69	8.8	89.6	
L1680-01	CONCRETE-WALL	19	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1681-01	SH-BORING-1	20	1.18	9.68	8.46	85.6	
L1681-02	SH-BORING-2	21	1.16	9.83	8.54	85.1	
L1682-01	S-1	35	1.18	9.57	8.76	90.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$