


LB107968

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-82	0.000	-82				
0.2	0.200	0.182	-0.018	574	0.0 %	574				
2.5	2.500	2.467	-0.033	8201	0.0 %	8201				
5.0	5.000	4.981	-0.019	16593	0.0 %	16593				
7.5	7.500	7.844	0.344	26152	0.0 %	26152				
10.0	10.000	9.741	-0.259	32484	0.0 %	32484				

LB107968 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Feb 2020 09:35:33				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-82	-	0.00007470A	PPB	26 Feb 2020 09:38:34	S	Std
0.2 - 1	50.2	574	-	0.20007470A	PPB	26 Feb 2020 09:40:42	S	Std
2.5 - 1	52.5	8201	-	2.50007470A	PPB	26 Feb 2020 09:42:49	S	Std
5.0 - 1	55	16593	-	5.00007470A	PPB	26 Feb 2020 09:44:56	S	Std
7.5 - 1	57.5	26152	-	7.50007470A	PPB	26 Feb 2020 09:47:04	S	Std
10.0 - 1	510	32484	-	10.00007470A	PPB	26 Feb 2020 09:51:42	S	Std
ICV33 - 1	ICV33	12780	3.8384	-7470A	PPB	26 Feb 2020 09:54:19	U	SMPL
ICB33 - 1	ICB33	-130	-0.0292	-7470A	PPB	26 Feb 2020 09:59:00	U	SMPL
CCV01 - 1	CCV01	16460	4.9409	-7470A	PPB	26 Feb 2020 10:01:06	U	SMPL
CCB01 - 1	CCB01	-334	-0.0903	-7470A	PPB	26 Feb 2020 10:03:11	U	SMPL
CRA - 1	CRA	499	0.1593	-7470A	PPB	26 Feb 2020 10:05:17	U	SMPL
HighStd - 1	HighStd	32335	9.6967	-7470A	PPB	26 Feb 2020 10:07:23	U	SMPL
ChkStd - 1	ChkStd	23861	7.1580	-7470A	PPB	26 Feb 2020 10:09:28	U	SMPL
PB127071BL - 1	PBW	-110	-0.0232	-7470A	PPB	26 Feb 2020 10:14:04	U	SMPL
PB127071BS - 1	LCSW	12686	3.8102	-7470A	PPB	26 Feb 2020 10:16:10	U	SMPL
L1367-08 - 1	OK-02-022420	-267	-0.0702	-7470A	PPB	26 Feb 2020 10:18:17	U	SMPL
L1367-08DUP - 1	OK-02-022420DUP	-90	-0.0172	-7470A	PPB	26 Feb 2020 10:20:24	U	SMPL
L1367-08MS - 1	OK-02-022420MS	11073	3.3270	-7470A	PPB	26 Feb 2020 10:22:31	U	SMPL
L1367-08MSD - 1	OK-02-022420MSD	11656	3.5017	-7470A	PPB	26 Feb 2020 10:24:39	U	SMPL
L1668-01 - 1	30875	-254	-0.0663	-7470A	PPB	26 Feb 2020 10:26:46	U	SMPL
CCV02 - 1	CCV02	17020	5.1086	-7470A	PPB	26 Feb 2020 10:28:53	U	SMPL
CCB02 - 1	CCB02	-283	-0.0750	-7470A	PPB	26 Feb 2020 10:30:58	U	SMPL
L1668-05 - 1	30830	1142	0.3519	-7470A	PPB	26 Feb 2020 10:33:05	U	SMPL
L1668-07 - 1	30793	168	0.0601	-7470A	PPB	26 Feb 2020 10:35:11	U	SMPL
L1669-01 - 1	VNJ-231	-9	0.0071	-7470A	PPB	26 Feb 2020 10:37:17	U	SMPL
L1669-03 - 1	72-11991	22	0.0164	-7470A	PPB	26 Feb 2020 10:39:24	U	SMPL
L1669-05 - 1	RBR-2520699	15	0.0143	-7470A	PPB	26 Feb 2020 10:41:30	U	SMPL
L1670-11 - 1	SOLID-COMP	4	0.0110	-7470A	PPB	26 Feb 2020 10:43:37	U	SMPL
PB127072BL - 1	PBW	42	0.0224	-7470A	PPB	26 Feb 2020 10:45:44	U	SMPL
PB127072BS - 1	LCSW	11150	3.3501	-7470A	PPB	26 Feb 2020 10:47:51	U	SMPL
L1666-01 - 1	NB-21944	-29	0.0011	-7470A	PPB	26 Feb 2020 10:49:59	U	SMPL
L1670-13 - 1	WATER-COMP	-86	-0.0160	-7470A	PPB	26 Feb 2020 10:52:05	U	SMPL
CCV03 - 1	CCV03	16608	4.9852	-7470A	PPB	26 Feb 2020 10:54:11	U	SMPL
CCB03 - 1	CCB03	-290	-0.0771	-7470A	PPB	26 Feb 2020 10:56:18	U	SMPL
L1670-13DUP - 1	WATER-COMPDUP	-109	-0.0229	-7470A	PPB	26 Feb 2020 10:58:25	U	SMPL
L1670-13MS - 1	WATER-COMPMS	12298	3.6940	-7470A	PPB	26 Feb 2020 11:00:31	U	SMPL
L1670-13MSD - 1	WATER-COMPMSD	12182	3.6593	-7470A	PPB	26 Feb 2020 11:02:37	U	SMPL
PB127073BL - 1	PBW	-256	-0.0669	-7470A	PPB	26 Feb 2020 11:04:44	U	SMPL
PB127073BS - 1	LCSW	11646	3.4987	-7470A	PPB	26 Feb 2020 11:06:51	U	SMPL
L1161-07 - 1	LOD-MDL-WATER-01-QT1-2020	146	0.0535	-7470A	PPB	26 Feb 2020 11:13:50	U	SMPL
L1161-08 - 1	LOQ-WATER-02-QT1-2020	549	0.1743	-7470A	PPB	26 Feb 2020 11:15:57	U	SMPL
L1671-02 - 1	FIELD-BLANK	-26	0.0020	-7470A	PPB	26 Feb 2020 11:18:05	U	SMPL
L1671-03 - 1	E156-B4-022420-GW	2409	0.7315	-7470A	PPB	26 Feb 2020 11:24:49	U	SMPL
L1671-03DUP - 1	E156-B4-022420-GWDUP	2023	0.6158	-7470A	PPB	26 Feb 2020 11:26:55	U	SMPL
CCV04 - 1	CCV04	18210	5.4651	-7470A	PPB	26 Feb 2020 11:29:01	U	SMPL
CCB04 - 1	CCB04	-24	0.0026	-7470A	PPB	26 Feb 2020 11:31:07	U	SMPL
L1671-03MS - 1	E156-B4-022420-GWMS	16100	4.8330	-7470A	PPB	26 Feb 2020 11:35:43	U	SMPL
L1671-03MSD - 1	E156-B4-022420-GWMSD	16088	4.8294	-7470A	PPB	26 Feb 2020 11:37:49	U	SMPL
L1671-04 - 1	E156-B4-022420-GW-BD	1589	0.4858	-7470A	PPB	26 Feb 2020 11:39:55	U	SMPL
L1367-08LX5 - 1		14	0.0140	-7470A	PPB	26 Feb 2020 11:42:01	U	SMPL
L1367-08A - 1		14104	4.2351	-7470A	PPB	26 Feb 2020 11:44:08	U	SMPL
L1670-13LX5 - 1		-15	0.0053	-7470A	PPB	26 Feb 2020 11:46:15	U	SMPL
L1670-13A - 1		13780	4.1380	-7470A	PPB	26 Feb 2020 11:48:22	U	SMPL
L1671-03LX5 - 1		356	0.1164	-7470A	PPB	26 Feb 2020 11:55:30	U	SMPL
L1671-03A - 1		16397	4.9220	-7470A	PPB	26 Feb 2020 11:57:37	U	SMPL
CCV05 - 1	CCV05	16736	5.0235	-7470A	PPB	26 Feb 2020 12:06:19	U	SMPL
CCB05 - 1	CCB05	-69	-0.0109	-7470A	PPB	26 Feb 2020 12:08:25	U	SMPL
PB127072TB - 1	PB127054TB	79	0.0335	-7470A	PPB	26 Feb 2020 12:10:32	U	SMPL
PB127072TB - 1	PB127068TB	51	0.0251	-7470A	PPB	26 Feb 2020 12:12:38	U	SMPL
CCV06 - 1	CCV06	16456	4.9397	-7470A	PPB	26 Feb 2020 12:14:44	U	SMPL
CCB06 - 1	CCB06	-93	-0.0181	-7470A	PPB	26 Feb 2020 12:16:50	U	SMPL