



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107970

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/25/2020	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/25/2020	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	02/25/2020	17:12
4	ICV	1	Water	NA	NA	20.1	6.99	02/25/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	02/25/2020	17:16
6	L1370-10	1	Solid	20.03	20	21.0	8.76	02/25/2020	17:19
7	L1673-01	1	Solid	20.04	20	21.8	9.49	02/25/2020	17:20
8	L1676-01	1	Solid	20.03	20	20.8	6.97	02/25/2020	17:22
9	L1676-03	1	Solid	20.01	20	20.7	6.61	02/25/2020	17:25
10	L1676-07	1	Solid	20.03	20	20.8	7.04	02/25/2020	17:30
11	L1678-02	1	Solid	20.03	20	21.5	8.26	02/25/2020	17:32
12	L1678-02DUP	1	Solid	20.04	20	21.6	8.27	02/25/2020	17:34
13	CCV2	1	Water	NA	NA	20.2	12.02	02/25/2020	17:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1370

WorkList ID : 136462

Department : Wet-Chemistry

Date : 02-25-2020 14:56:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/25/2020	Solid	L1370-10	Corrosivity	Cool 4 deg C	PSEG05		OR-03-022520	02/25/2020	9045D
02/24/2020	Solid	L1673-01	Corrosivity	Cool 4 deg C	PSEG01	I33	WH-COMP-3	02/24/2020	9045D
02/25/2020	Solid	L1676-01	Corrosivity	Cool 4 deg C	PSEG01	I41	DRUMS-1-2	02/25/2020	9045D
02/25/2020	Solid	L1676-03	Corrosivity	Cool 4 deg C	PSEG01	I41	DRUMS-3-4	02/25/2020	9045D
02/25/2020	Solid	L1676-07	Corrosivity	Cool 4 deg C	PSEG01	I41	RO-1-2	02/25/2020	9045D
02/25/2020	Solid	L1678-02	Corrosivity	Cool 4 deg C	PSEG01	I41	VNJ-241	02/25/2020	9045D

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: