

LB 107979


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50765



A= 0.0000e+000

B= 1.9367e-004

C= -9.7108e-003

Rho= 0.9992682

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-49	0.000	-49				
0.2	0.200	0.190	-0.010	1032	0.0 %	1032				
2.5	2.500	2.527	0.027	13097	0.0 %	13097				
5.0	5.000	4.903	-0.097	25364	0.0 %	25364				
7.5	7.500	7.778	0.278	40210	0.0 %	40210				
10.0	10.000	9.822	-0.178	50765	0.0 %	50765				

LB107979 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 26 Feb 2020 15:42:18					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-49	-	0.00007471B	PPB		26 Feb 2020 15:43:56	S	Std
0.2 - 1	50.2	1032	-	0.20007471B	PPB		26 Feb 2020 15:46:13	S	Std
2.5 - 1	52.5	13097	-	2.50007471B	PPB		26 Feb 2020 15:48:30	S	Std
5.0 - 1	55	25364	-	5.00007471B	PPB		26 Feb 2020 15:50:46	S	Std
7.5 - 1	57.5	40210	-	7.50007471B	PPB		26 Feb 2020 15:53:03	S	Std
10.0 - 1	510	50765	-	10.00007471B	PPB		26 Feb 2020 15:58:01	S	Std
ICV34 - 1	ICV34	19089	3.6873	-7471B	PPB		26 Feb 2020 16:00:40	U	SMPL
ICB34 - 1	ICB34	-348	-0.0771	-7471B	PPB		26 Feb 2020 16:02:55	U	SMPL
CCV10 - 1	CCV10	25338	4.8975	-7471B	PPB		26 Feb 2020 16:05:10	U	SMPL
CCB10 - 1	CCB10	-339	-0.0754	-7471B	PPB		26 Feb 2020 16:07:25	U	SMPL
CRA - 1	CRA	991	0.1822	-7471B	PPB		26 Feb 2020 16:09:42	U	SMPL
HighStd - 1	HighStd	51004	9.8682	-7471B	PPB		26 Feb 2020 16:14:28	U	SMPL
ChkStd - 1	ChkStd	39477	7.6358	-7471B	PPB		26 Feb 2020 16:16:43	U	SMPL
PB127084BL - 1	PBS	-169	-0.0424	-7471B	PPB		26 Feb 2020 16:19:01	U	SMPL
L1161-03 - 1	MDL SOIL-03-QT1-2020	-14	-0.0424	-7471B	PPB		26 Feb 2020 16:23:34	U	SMPL
L1370-06 - 1	OR-03-022520	4669	0.8945	-7471B	PPB		26 Feb 2020 16:25:53	U	SMPL
L1675-01 - 1	COMP-FRONT-PARKING	22855	4.4166	-7471B	PPB		26 Feb 2020 16:28:10	U	SMPL
L1676-05 - 1	RO-1-2	56	0.0011	-7471B	PPB		26 Feb 2020 16:30:27	U	SMPL
L1678-01 - 1	VNJ-241	2973	0.5661	-7471B	PPB		26 Feb 2020 16:32:42	U	SMPL
CCV11 - 1	CCV11	25578	4.9440	-7471B	PPB		26 Feb 2020 16:34:58	U	SMPL
CCB11 - 1	CCB11	-107	-0.0304	-7471B	PPB		26 Feb 2020 16:37:13	U	SMPL
L1681-01 - 1	SH-BORING-1	1932	0.3645	-7471B	PPB		26 Feb 2020 16:39:29	U	SMPL
L1681-02 - 1	SH-BORING-2	1330	0.2479	-7471B	PPB		26 Feb 2020 16:46:55	U	SMPL
L1681-02DUP - 1	SH-BORING-2DUP	1198	0.2223	-7471B	PPB		26 Feb 2020 16:49:10	U	SMPL
L1681-02MS - 1	SH-BORING-2MS	21543	4.1625	-7471B	PPB		26 Feb 2020 16:51:26	U	SMPL
L1681-02MSD - 1	SH-BORING-2MSD	21793	4.2109	-7471B	PPB		26 Feb 2020 16:53:42	U	SMPL
L1682-01 - 1	S-1	2289	0.4336	-7471B	PPB		26 Feb 2020 16:55:58	U	SMPL
L1681-02LX5 - 1		241	0.0370	-7471B	PPB		26 Feb 2020 16:58:14	U	SMPL
L1681-02A - 1		21806	4.2135	-7471B	PPB		26 Feb 2020 17:00:31	U	SMPL
L1161-03 - 1	MDL-SOIL-03-QT	357	0.0594	-7471B	PPB		26 Feb 2020 17:05:20	U	SMPL
LB127084BS - 1	LCSS	19272	3.7227	-7471B	PPB		26 Feb 2020 17:15:16	U	SMPL
CCV12 - 1	CCV12	25481	4.9252	-7471B	PPB		26 Feb 2020 17:17:31	U	SMPL
CCB12 - 1	CCB12	-216	-0.0515	-7471B	PPB		26 Feb 2020 17:19:47	U	SMPL