



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107984

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	02/26/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	02/26/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	02/26/2020	17:19
4	ICV	1	Water	NA	NA	20.2	7.00	02/26/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.01	02/26/2020	17:22
6	L1675-01	1	Solid	20.03	20	20.5	7.85	02/26/2020	17:25
7	L1675-01DUP	1	Solid	20.04	20	20.6	7.86	02/26/2020	17:26
8	L1701-01	1	Solid	20.02	20	20.7	8.35	02/26/2020	17:29
9	CCV2	1	Water	NA	NA	20.1	12.02	02/26/2020	17:30

WORKLIST(Hardcopy Internal Chain)

13107984

WorkList Name : PH L1701 WorkList ID : 136500 Department : Wet-Chemistry Date : 02-26-2020 17:07:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/25/2020	Solid	L1675-01	pH	Cool 4 deg C	LAFO01	L22	COMP-FRONT-PARKING	02/25/2020	9045D
02/26/2020	Solid	L1701-01	pH	Cool 4 deg C	PSEG01	O12	N48082	02/26/2020	9045D

Date/Time 02-28-2020 17:10
 Received by: JN
 Relinquished by: JN

Date/Time 02-28-2020 18:20
 Received by: JN
 Relinquished by: JN