

LB108009


7471B

INSTRUMENT ID: CX1

Linear

 μ Abs.:

52153



A= 0.0000e+000

B= 1.9277e-004

C= -8.3377e-002

Rho= 0.9999361

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	510	0.000	510				
0.2	0.200	0.216	0.016	1553	0.0 %	1553				
2.5	2.500	2.422	-0.078	12996	0.0 %	12996				
5.0	5.000	5.039	0.039	26573	0.0 %	26573				
7.5	7.500	7.538	0.038	39536	0.0 %	39536				
10.0	10.000	9.970	-0.030	52153	0.0 %	52153				

LB108009 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Feb 2020 09:06:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	510	-	0.0000	7471B	PPB	28 Feb 2020 09:18:37	S	Std
0.2 - 1	50.2	1553	-	0.2000	7471B	PPB	28 Feb 2020 09:20:53	S	Std
2.5 - 1	52.5	12996	-	2.5000	7471B	PPB	28 Feb 2020 09:23:10	S	Std
5.0 - 1	55	26573	-	5.0000	7471B	PPB	28 Feb 2020 09:25:27	S	Std
7.5 - 1	57.5	39536	-	7.5000	7471B	PPB	28 Feb 2020 09:27:43	S	Std
10.0 - 1	510	52153	-	10.0000	7471B	PPB	28 Feb 2020 09:32:46	S	Std
ICV37 - 1	ICV37	19928	3.7581	-	7471B	PPB	28 Feb 2020 09:35:22	U	SMPL
ICB37 - 1	ICB37	134	-0.0575	-	7471B	PPB	28 Feb 2020 09:37:39	U	SMPL
CCV21 - 1	CCV21	27867	5.2885	-	7471B	PPB	28 Feb 2020 09:39:56	U	SMPL
CCB21 - 1	CCB21	148	-0.0548	-	7471B	PPB	28 Feb 2020 09:42:10	U	SMPL
CRA - 1	CRA	1475	0.2010	-	7471B	PPB	28 Feb 2020 09:44:27	U	SMPL
HighStd - 1	HighStd	52720	10.0794	-	7471B	PPB	28 Feb 2020 09:46:42	U	SMPL
ChkStd - 1	ChkStd	38658	7.3687	-	7471B	PPB	28 Feb 2020 09:48:57	U	SMPL
PB127134BL - 1	PBS	162	-0.0521	-	7471B	PPB	28 Feb 2020 09:51:15	U	SMPL
PB127134BS - 1	LCSS	20580	3.8838	-	7471B	PPB	28 Feb 2020 09:53:33	U	SMPL
L1373-04 - 1	TR-06-022720	4558	0.7953	-	7471B	PPB	28 Feb 2020 09:55:49	U	SMPL
L1705-01 - 1	FLY-ASH	9298	1.7090	-	7471B	PPB	28 Feb 2020 09:58:07	U	SMPL
L1705-02 - 1	BRICK	515	0.0159	-	7471B	PPB	28 Feb 2020 10:00:24	U	SMPL
L1705-03 - 1	BRICK-MORTAR	3110	0.5161	-	7471B	PPB	28 Feb 2020 10:02:42	U	SMPL
L1718-01 - 1	72-11976	3747	0.6389	-	7471B	PPB	28 Feb 2020 10:04:58	U	SMPL
CCV22 - 1	CCV22	28084	5.3304	-	7471B	PPB	28 Feb 2020 10:07:13	U	SMPL
CCB22 - 1	CCB22	404	-0.0055	-	7471B	PPB	28 Feb 2020 10:09:28	U	SMPL
L1718-04 - 1	72-11986	2869	0.4697	-	7471B	PPB	28 Feb 2020 10:11:45	U	SMPL
L1718-07 - 1	VNJ-232	43285	8.2607	-	7471B	PPB	28 Feb 2020 10:14:00	U	SMPL
L1718-10 - 1	O6	3823	0.6536	-	7471B	PPB	28 Feb 2020 10:16:15	U	SMPL
L1718-10DUP - 1	O6DUP	3730	0.6357	-	7471B	PPB	28 Feb 2020 10:18:32	U	SMPL
L1718-10MS - 1	O6MS	23726	4.4903	-	7471B	PPB	28 Feb 2020 10:20:48	U	SMPL
L1718-10MSD - 1	O6MSD	23533	4.4531	-	7471B	PPB	28 Feb 2020 10:23:05	U	SMPL
L1719-01 - 1	MP-1-COMP	1854	0.2740	-	7471B	PPB	28 Feb 2020 10:25:21	U	SMPL
L1718-10LX5 - 1		1152	0.1387	-	7471B	PPB	28 Feb 2020 10:27:38	U	SMPL
L1718-10A - 1		26185	4.9643	-	7471B	PPB	28 Feb 2020 10:29:55	U	SMPL
CCV23 - 1	CCV23	28418	5.3948	-	7471B	PPB	28 Feb 2020 10:32:11	U	SMPL
CCB23 - 1	CCB23	422	-0.0020	-	7471B	PPB	28 Feb 2020 10:34:26	U	SMPL