


LB 108015

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	-187	0.000	-187				
0.05	0.050	0.058	0.008	124	0.0 %	124				
0.2	0.200	0.205	0.005	648	0.0 %	648				
2.5	2.500	2.548	0.048	9027	0.0 %	9027				
5.0	5.000	5.058	0.058	17998	0.0 %	17998				
7.5	7.500	7.324	-0.176	26100	0.0 %	26100				
10.0	10.000	10.087	0.087	35978	0.0 %	35978				

LB108015 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Feb 2020 11:22:22

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-187	-	0.0000245.1	PPB	28 Feb 2020 11:56:15	S	Std
0.05 - 1	50.05	124	-	0.0500245.1	PPB	28 Feb 2020 11:58:32	S	Std
0.2 - 1	50.2	648	-	0.2000245.1	PPB	28 Feb 2020 12:00:49	S	Std
2.5 - 1	52.5	9027	-	2.5000245.1	PPB	28 Feb 2020 12:03:06	S	Std
5.0 - 1	55	17998	-	5.0000245.1	PPB	28 Feb 2020 12:05:22	S	Std
7.5 - 1	57.5	26100	-	7.5000245.1	PPB	28 Feb 2020 12:07:38	S	Std
10.0 - 1	510	35978	-	10.0000245.1	PPB	28 Feb 2020 12:09:55	S	Std
ICV38 - 1	ICV38	13753	3.8702	-245.1	PPB	28 Feb 2020 12:15:18	U	SMPL
ICB38 - 1	ICB38	-221	-0.0385	-245.1	PPB	28 Feb 2020 12:17:32	U	SMPL
CCV24 - 1	CCV24	17971	5.0500	-245.1	PPB	28 Feb 2020 12:19:48	U	SMPL
CCB24 - 1	CCB24	-176	-0.0259	-245.1	PPB	28 Feb 2020 12:22:03	U	SMPL
CRA - 1	CRA	97	0.0505	-245.1	PPB	28 Feb 2020 12:26:54	U	SMPL
HighStd - 1	HighStd	36028	10.1007	-245.1	PPB	28 Feb 2020 12:29:09	U	SMPL
ChkStd - 1	ChkStd	28334	7.9486	-245.1	PPB	28 Feb 2020 12:31:24	U	SMPL
PB127146BL - 1	PBW	-293	-0.0586	-245.1	PPB	28 Feb 2020 12:37:56	U	SMPL
PB127146BS - 1	LCSW	14049	3.9530	-245.1	PPB	28 Feb 2020 12:42:47	U	SMPL
L1161-10 - 1	LOD-MDL-WATER-04-QT1-2020	121	0.0572	-245.1	PPB	28 Feb 2020 12:47:38	U	SMPL
L1161-11 - 1	LOQ-WATER-05-QT1-2020	646	0.2041	-245.1	PPB	28 Feb 2020 12:49:54	U	SMPL
L1699-07 - 1	MH-1-COMP-2-27-2020	667	0.2099	-245.1	PPB	28 Feb 2020 13:04:40	U	SMPL
L1699-07DUP - 1	MH-1-COMP-2-27-2020DUP	625	0.1982	-245.1	PPB	28 Feb 2020 13:06:57	U	SMPL
L1699-07MS - 1	MH-1-COMP-2-27-2020MS	16398	4.6100	-245.1	PPB	28 Feb 2020 13:09:12	U	SMPL
CCV25 - 1	CCV25	16891	4.7479	-245.1	PPB	28 Feb 2020 13:27:07	U	SMPL
CCB25 - 1	CCB25	-24	0.0167	-245.1	PPB	28 Feb 2020 13:29:23	U	SMPL
L1699-07MSD - 1	MH-1-COMP-2-27-2020MSD	15115	4.2512	-245.1	PPB	28 Feb 2020 13:34:17	U	SMPL
L1699-07LX5 - 1		136	0.0614	-245.1	PPB	28 Feb 2020 13:36:32	U	SMPL
L1699-07A - 1		16827	4.7300	-245.1	PPB	28 Feb 2020 13:38:47	U	SMPL
CCV26 - 1	CCV26	17958	5.0464	-245.1	PPB	28 Feb 2020 13:41:03	U	SMPL
CCB26 - 1	CCB26	13	0.0270	-245.1	PPB	28 Feb 2020 13:43:19	U	SMPL