

LB 108021


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36465



A= 0.0000e+000

B= 2.7403e-004

C= 2.7527e-002

Rho= 0.9998623

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	13	0.000	13				
0.2	0.200	0.211	0.011	670	0.0 %	670				
2.5	2.500	2.406	-0.094	8678	0.0 %	8678				
5.0	5.000	5.092	0.092	18482	0.0 %	18482				
7.5	7.500	7.440	-0.060	27050	0.0 %	27050				
10.0	10.000	10.020	0.020	36465	0.0 %	36465				

LB108021 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 28 Feb 2020 14:30:28							
Sample ID	Extended ID	µ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	13	-	0.0000	7470A	PPB	28 Feb 2020 14:32:56	S	Std
0.2 - 1	50.2	670	-	0.2000	7470A	PPB	28 Feb 2020 14:35:03	S	Std
2.5 - 1	52.5	8678	-	2.5000	7470A	PPB	28 Feb 2020 14:37:11	S	Std
5.0 - 1	55	18482	-	5.0000	7470A	PPB	28 Feb 2020 14:39:18	S	Std
7.5 - 1	57.5	27050	-	7.5000	7470A	PPB	28 Feb 2020 14:41:26	S	Std
10.0 - 1	510	36465	-	10.0000	7470A	PPB	28 Feb 2020 14:43:33	S	Std
ICV39 - 1	ICV39	13277	3.6658	-7470A	PPB	28 Feb 2020 14:46:08	U	SMPL	
ICB39 - 1	ICB39	-264	-0.0448	-7470A	PPB	28 Feb 2020 14:48:13	U	SMPL	
CCV27 - 1	CCV27	18109	4.9899	-7470A	PPB	28 Feb 2020 14:50:19	U	SMPL	
CCB27 - 1	CCB27	-283	-0.0500	-7470A	PPB	28 Feb 2020 14:52:25	U	SMPL	
CRA - 1	CRA	588	0.1887	-7470A	PPB	28 Feb 2020 14:54:31	U	SMPL	
HighStd - 1	HighStd	36199	9.9471	-7470A	PPB	28 Feb 2020 14:56:36	U	SMPL	
ChkStd - 1	ChkStd	26066	7.1704	-7470A	PPB	28 Feb 2020 14:58:42	U	SMPL	
PB127154BL - 1	PBW	-401	-0.0824	-7470A	PPB	28 Feb 2020 15:00:51	U	SMPL	
PB127154BS - 1	LCSW	14655	4.0434	-7470A	PPB	28 Feb 2020 15:02:58	U	SMPL	
L1363-10 - 1	HR-05-022620	-261	-0.0440	-7470A	PPB	28 Feb 2020 15:05:05	U	SMPL	
L1363-10DUP - 1	HR-05-022620DUP	-70	0.0083	-7470A	PPB	28 Feb 2020 15:07:12	U	SMPL	
L1363-10MS - 1	HR-05-022620MS	12954	3.5773	-7470A	PPB	28 Feb 2020 15:09:20	U	SMPL	
L1363-10MSD - 1	HR-05-022620MSD	13879	3.8308	-7470A	PPB	28 Feb 2020 15:11:28	U	SMPL	
L1373-06 - 1	TR-06-022720	-217	-0.0319	-7470A	PPB	28 Feb 2020 15:13:35	U	SMPL	
CCV28 - 1	CCV28	18436	5.0795	-7470A	PPB	28 Feb 2020 15:15:41	U	SMPL	
CCB28 - 1	CCB28	-261	-0.0440	-7470A	PPB	28 Feb 2020 15:17:47	U	SMPL	
L1581-04 - 1	40712	-71	0.0081	-7470A	PPB	28 Feb 2020 15:19:54	U	SMPL	
L1718-03 - 1	72-11976	-30	0.0193	-7470A	PPB	28 Feb 2020 15:22:00	U	SMPL	
L1718-06 - 1	72-11986	5	0.0289	-7470A	PPB	28 Feb 2020 15:24:06	U	SMPL	
L1718-09 - 1	VNJ-232	1	0.0278	-7470A	PPB	28 Feb 2020 15:26:13	U	SMPL	
L1718-12 - 1	06	-3	0.0267	-7470A	PPB	28 Feb 2020 15:28:19	U	SMPL	
L1719-05 - 1	MP-1-COMP	-4	0.0264	-7470A	PPB	28 Feb 2020 15:30:26	U	SMPL	
PB127155BL - 1	PBW	1	0.0278	-7470A	PPB	28 Feb 2020 15:32:33	U	SMPL	
PB127155BS - 1	LCSW	14758	4.0717	-7470A	PPB	28 Feb 2020 15:34:40	U	SMPL	
L1161-09 - 1	MDL-WATER-03-QT1-2020	110	0.0577	-7470A	PPB	28 Feb 2020 15:41:44	U	SMPL	
L1683-27 - 1	FB-02-20200225	0	0.0275	-7470A	PPB	28 Feb 2020 15:43:49	U	SMPL	
CCV29 - 1	CCV29	18165	5.0053	-7470A	PPB	28 Feb 2020 15:45:55	U	SMPL	
CCB29 - 1	CCB29	-214	-0.0311	-7470A	PPB	28 Feb 2020 15:48:02	U	SMPL	
L1717-01 - 1	PLANK-MANHOLE	335	0.1193	-7470A	PPB	28 Feb 2020 15:50:09	U	SMPL	
L1717-01DUP - 1	PLANK-MANHOLEDUP	410	0.1399	-7470A	PPB	28 Feb 2020 15:52:15	U	SMPL	
L1717-01MS - 1	PLANK-MANHOLEMS	11769	3.2526	-7470A	PPB	28 Feb 2020 15:54:21	U	SMPL	
L1717-01MSD - 1	PLANK-MANHOLEMSD	11945	3.3008	-7470A	PPB	28 Feb 2020 15:56:27	U	SMPL	
L1363-10LX5 - 1		-113	-0.0034	-7470A	PPB	28 Feb 2020 15:58:35	U	SMPL	
L1363-10A - 1		15463	4.2649	-7470A	PPB	28 Feb 2020 16:00:43	U	SMPL	
L1717-01LX5 - 1		-53	0.0130	-7470A	PPB	28 Feb 2020 16:02:50	U	SMPL	
L1717-01A - 1		15113	4.1689	-7470A	PPB	28 Feb 2020 16:04:57	U	SMPL	
PB127154TB - 1	PB127122TB	-42	0.0160	-7470A	PPB	28 Feb 2020 16:07:05	U	SMPL	
CCV30 - 1	CCV30	18584	5.1201	-7470A	PPB	28 Feb 2020 16:09:11	U	SMPL	
CCB30 - 1	CCB30	-193	-0.0254	-7470A	PPB	28 Feb 2020 16:11:17	U	SMPL	