



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108030

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	02/28/2020	14:00
2	CAL2	1	Water	NA	NA	20.1	7.00	02/28/2020	14:01
3	CAL3	1	Water	NA	NA	20.2	10.03	02/28/2020	14:02
4	ICV	1	Water	NA	NA	20.1	6.99	02/28/2020	14:03
5	CCV1	1	Water	NA	NA	20.2	2.02	02/28/2020	17:00
6	L1369-06	1	Solid	20.03	20	20.5	8.57	02/28/2020	17:02
7	L1372-10	1	Solid	20.04	20	19.7	8.95	02/28/2020	17:05
8	L1745-02	1	Solid	20.02	20	20.2	6.84	02/28/2020	17:06
9	L1745-02DUP	1	Solid	20.03	20	20.3	6.85	02/28/2020	17:08
10	CCV2	1	Water	NA	NA	20.1	12.02	02/28/2020	17:10

V0108030

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1372 WorkList ID : 136578 Department : Wet-Chemistry Date : 02-28-2020 14:26:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/28/2020	Solid	L1369-06	Corrosivity	Cool 4 deg C	PSEG05	I42	PL-02-022820	02/28/2020	9045D
02/28/2020	Solid	L1372-10	Corrosivity	Cool 4 deg C	PSEG05		SU-03-022820	02/28/2020	9045D
02/28/2020	Solid	L1745-02	Corrosivity	Cool 4 deg C	PSEG01	I42	MP-2-COMP	02/28/2020	9045D

Date/Time 02-28-2020 16:20
Received by: JP
Relinquished by: JP

Date/Time 02-28-2020 18:30
Received by: JP
Relinquished by: JP