


LB 108037

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	-40	0.000	-40				
0.2	0.200	0.210	0.010	905	0.0 %	905				
2.5	2.500	2.473	-0.027	12976	0.0 %	12976				
5.0	5.000	4.901	-0.099	25921	0.0 %	25921				
7.5	7.500	7.622	0.122	40434	0.0 %	40434				
10.0	10.000	9.961	-0.039	52906	0.0 %	52906				

LB108037 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 02 Mar 2020 09:05:18					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-40	-	0.0000	7471B	PPB	02 Mar 2020 09:13:20	S	Std
0.2 - 1	50.2	905	-	0.2000	7471B	PPB	02 Mar 2020 09:15:37	S	Std
2.5 - 1	52.5	12976	-	2.5000	7471B	PPB	02 Mar 2020 09:17:54	S	Std
5.0 - 1	55	25921	-	5.0000	7471B	PPB	02 Mar 2020 09:20:11	S	Std
7.5 - 1	57.5	40434	-	7.5000	7471B	PPB	02 Mar 2020 09:22:27	S	Std
10.0 - 1	510	52906	-	10.0000	7471B	PPB	02 Mar 2020 09:24:43	S	Std
ICV40 - 1	ICV40	20105	3.8102	-	7471B	PPB	02 Mar 2020 09:27:36	U	SMPL
ICB40 - 1	ICB40	-267	-0.0102	-	7471B	PPB	02 Mar 2020 09:29:50	U	SMPL
CCV31 - 1	CCV31	27444	5.1865	-	7471B	PPB	02 Mar 2020 09:32:06	U	SMPL
CCB31 - 1	CCB31	-297	-0.0158	-	7471B	PPB	02 Mar 2020 09:34:21	U	SMPL
CRA - 1	CRA	916	0.2117	-	7471B	PPB	02 Mar 2020 09:36:38	U	SMPL
HighStd - 1	HighStd	53437	10.0609	-	7471B	PPB	02 Mar 2020 09:38:53	U	SMPL
ChkStd - 1	ChkStd	37801	7.1287	-	7471B	PPB	02 Mar 2020 09:41:08	U	SMPL
PB127184BL - 1	PBS	-436	-0.0419	-	7471B	PPB	02 Mar 2020 09:43:25	U	SMPL
PB127184BS - 1	LCSS	21188	4.0133	-	7471B	PPB	02 Mar 2020 09:45:43	U	SMPL
L1369-01 - 1	PL-02-022820	4240	0.8350	-	7471B	PPB	02 Mar 2020 09:47:59	U	SMPL
L1372-06 - 1	SU-03-022820	13974	2.6604	-	7471B	PPB	02 Mar 2020 09:50:17	U	SMPL
L1723-01 - 1	B3-(1.5-2)	10264	1.9647	-	7471B	PPB	02 Mar 2020 09:52:34	U	SMPL
L1723-02 - 1	B2-(1.5-2)	1979	0.4110	-	7471B	PPB	02 Mar 2020 09:54:51	U	SMPL
L1735-01 - 1	FK-SF-01-009-A	7579	1.4612	-	7471B	PPB	02 Mar 2020 09:57:06	U	SMPL
CCV32 - 1	CCV32	28222	5.3324	-	7471B	PPB	02 Mar 2020 09:59:22	U	SMPL
CCB32 - 1	CCB32	-58	0.0290	-	7471B	PPB	02 Mar 2020 10:01:37	U	SMPL
L1735-01DUP - 1	FK-SF-01-009-ADUP	7134	1.3777	-	7471B	PPB	02 Mar 2020 10:06:39	U	SMPL
L1735-01MS - 1	FK-SF-01-009-AMS	27059	5.1143	-	7471B	PPB	02 Mar 2020 10:08:54	U	SMPL
L1735-01MSD - 1	FK-SF-01-009-AMSD	28166	5.3219	-	7471B	PPB	02 Mar 2020 10:11:09	U	SMPL
L1735-03 - 1	FK-SF-01-009-B	6562	1.2705	-	7471B	PPB	02 Mar 2020 10:13:26	U	SMPL
L1736-01 - 1	FK-SF-01-010-A	7673	1.4788	-	7471B	PPB	02 Mar 2020 10:15:42	U	SMPL
L1736-03 - 1	FK-SF-01-010-B	4444	0.8733	-	7471B	PPB	02 Mar 2020 10:17:58	U	SMPL
L1737-01 - 1	FK-BPM-03-013-ABCD	1634	0.3463	-	7471B	PPB	02 Mar 2020 10:20:15	U	SMPL
L1738-01 - 1	FK-BPM-01-G0-G1-G0A-G1A	1145	0.2546	-	7471B	PPB	02 Mar 2020 10:22:31	U	SMPL
L1739-01 - 1	FK-BPM-04-007-ABCD	1120	0.2499	-	7471B	PPB	02 Mar 2020 10:24:49	U	SMPL
L1740-01 - 1	GP-06-022720	2067	0.4275	-	7471B	PPB	02 Mar 2020 10:27:04	U	SMPL
CCV33 - 1	CCV33	28114	5.3121	-	7471B	PPB	02 Mar 2020 10:29:20	U	SMPL
CCB33 - 1	CCB33	-62	0.0283	-	7471B	PPB	02 Mar 2020 10:31:35	U	SMPL
L1740-02 - 1	GP-07-022720	1129	0.2516	-	7471B	PPB	02 Mar 2020 10:33:51	U	SMPL
L1740-03 - 1	GP-08-022720	1681	0.3551	-	7471B	PPB	02 Mar 2020 10:36:06	U	SMPL
L1740-04 - 1	GP-09-022720	2575	0.5228	-	7471B	PPB	02 Mar 2020 10:38:22	U	SMPL
L1740-06 - 1	GP-12-022720	2341	0.4789	-	7471B	PPB	02 Mar 2020 10:40:37	U	SMPL
L1740-07 - 1	GP-13-022720	4661	0.9140	-	7471B	PPB	02 Mar 2020 10:42:52	U	SMPL
PB127185BL - 1	PBS	26	0.0448	-	7471B	PPB	02 Mar 2020 10:45:09	U	SMPL
PB127185BS - 1	LCSS	19665	3.7277	-	7471B	PPB	02 Mar 2020 10:47:25	U	SMPL
L1741-01 - 1	TP-1	1467	0.3150	-	7471B	PPB	02 Mar 2020 10:49:41	U	SMPL
L1741-01DUP - 1	TP-1DUP	1310	0.2856	-	7471B	PPB	02 Mar 2020 10:51:58	U	SMPL
L1741-01MS - 1	TP-1MS	21816	4.1311	-	7471B	PPB	02 Mar 2020 10:54:13	U	SMPL
CCV34 - 1	CCV34	28169	5.3224	-	7471B	PPB	02 Mar 2020 10:56:29	U	SMPL
CCB34 - 1	CCB34	-63	0.0281	-	7471B	PPB	02 Mar 2020 10:58:44	U	SMPL
L1741-01MSD - 1	TP-1MSD	24189	4.5761	-	7471B	PPB	02 Mar 2020 11:01:00	U	SMPL
L1741-03 - 1	TP-2	1920	0.3999	-	7471B	PPB	02 Mar 2020 11:03:16	U	SMPL
L1744-01 - 1	TP-3	1037	0.2344	-	7471B	PPB	02 Mar 2020 11:05:32	U	SMPL
L1744-02 - 1	TP-2	1176	0.2604	-	7471B	PPB	02 Mar 2020 11:07:47	U	SMPL
L1744-03 - 1	TP-1	915	0.2115	-	7471B	PPB	02 Mar 2020 11:10:03	U	SMPL
L1744-04 - 1	TP-4	863	0.2017	-	7471B	PPB	02 Mar 2020 11:12:19	U	SMPL
L1745-01 - 1	MP-2-COMP	323	0.1005	-	7471B	PPB	02 Mar 2020 11:14:36	U	SMPL
CCV35 - 1	CCV35	28048	5.2997	-	7471B	PPB	02 Mar 2020 11:16:53	U	SMPL
CCB35 - 1	CCB35	-36	0.0331	-	7471B	PPB	02 Mar 2020 11:19:10	U	SMPL
L1735-01LX5 - 1		1420	0.3062	-	7471B	PPB	02 Mar 2020 11:21:26	U	SMPL
L1735-01A - 1		28691	5.4203	-	7471B	PPB	02 Mar 2020 11:23:41	U	SMPL
L1741-01LX5 - 1		207	0.0787	-	7471B	PPB	02 Mar 2020 11:25:56	U	SMPL
L1741-01A - 1		23207	4.3919	-	7471B	PPB	02 Mar 2020 11:28:12	U	SMPL
CCV36 - 1	CCV36	28088	5.3072	-	7471B	PPB	02 Mar 2020 11:30:28	U	SMPL
CCB36 - 1	CCB36	-53	0.0300	-	7471B	PPB	02 Mar 2020 11:32:43	U	SMPL