



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108040

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/02/2020	11:45
2	CAL2	1	Water	NA	NA	20.1	7.01	03/02/2020	11:46
3	CAL3	1	Water	NA	NA	20.1	10.04	03/02/2020	11:47
4	ICV	1	Water	NA	NA	20.1	6.99	03/02/2020	11:49
5	CCV1	1	Water	NA	NA	20.2	2.01	03/02/2020	11:50
6	L1747-01	1	Water	NA	NA	21.5	5.41	03/02/2020	12:00
7	L1747-01DUP	1	Water	NA	NA	21.5	5.42	03/02/2020	12:02
8	CCV2	1	Water	NA	NA	20.1	12.02	03/02/2020	12:05

W3108046

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1747 WorkList ID : 136621 Department : Wet-Chemistry Date : 03-02-2020 11:35:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2020	Water	L1747-01- 1	pH	Cool 4 deg C	STER08	O11	POOL-WATER	02/27/2020	SM 4500-H B

Date/Time 03-02-2020 11:40
Received by: JP
Relinquished by: JP

Date/Time 03-02-2020 12:50
Received by: JP
Relinquished by: JP