


LB108048

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	18	0.000	18				
0.2	0.200	0.141	-0.059	564	0.0 %	564				
2.5	2.500	2.439	-0.061	9292	0.0 %	9292				
5.0	5.000	5.242	0.242	19934	0.0 %	19934				
7.5	7.500	7.472	-0.028	28406	0.0 %	28406				
10.0	10.000	9.909	-0.091	37658	0.0 %	37658				

LB108048 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 02 Mar 2020 14:40:38					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	18	-	0.0000	7470A	PPB	02 Mar 2020 14:43:22	S	Std
0.2 - 1	S0.2	564	-	0.2000	7470A	PPB	02 Mar 2020 14:45:30	S	Std
2.5 - 1	S2.5	9292	-	2.5000	7470A	PPB	02 Mar 2020 14:47:37	S	Std
5.0 - 1	S5	19934	-	5.0000	7470A	PPB	02 Mar 2020 14:49:44	S	Std
7.5 - 1	S7.5	28406	-	7.5000	7470A	PPB	02 Mar 2020 14:51:51	S	Std
10.0 - 1	S10	37658	-	10.0000	7470A	PPB	02 Mar 2020 14:56:38	S	Std
ICV41 - 1	ICV41	13726	3.6068	-	7470A	PPB	02 Mar 2020 14:59:15	U	SMPL
ICB41 - 1	ICB41	-303	-0.0874	-	7470A	PPB	02 Mar 2020 15:01:20	U	SMPL
CCV37 - 1	CCV37	19170	5.0404	-	7470A	PPB	02 Mar 2020 15:03:27	U	SMPL
CCB37 - 1	CCB37	-291	-0.0843	-	7470A	PPB	02 Mar 2020 15:05:32	U	SMPL
CRA - 1	CRA	578	0.1446	-	7470A	PPB	02 Mar 2020 15:10:04	U	SMPL
HighStd - 1	HighStd	36968	9.7271	-	7470A	PPB	02 Mar 2020 15:12:09	U	SMPL
ChkStd - 1	ChkStd	25502	6.7078	-	7470A	PPB	02 Mar 2020 15:14:15	U	SMPL
PB127209BL - 1	PBW	-107	-0.0358	-	7470A	PPB	02 Mar 2020 15:18:51	U	SMPL
PB127209BS - 1	LCSW	14350	3.7711	-	7470A	PPB	02 Mar 2020 15:20:58	U	SMPL
L1369-06 - 1	PL-02-022820	-234	-0.0693	-	7470A	PPB	02 Mar 2020 15:23:05	U	SMPL
L1372-10 - 1	SU-03-022820	-75	-0.0274	-	7470A	PPB	02 Mar 2020 15:25:13	U	SMPL
L1372-10DUP - 1	SU-03-022820DUP	-54	-0.0219	-	7470A	PPB	02 Mar 2020 15:27:20	U	SMPL
L1372-10MS - 1	SU-03-022820MS	13823	3.6324	-	7470A	PPB	02 Mar 2020 15:29:28	U	SMPL
L1372-10MSD - 1	SU-03-022820MSD	13904	3.6537	-	7470A	PPB	02 Mar 2020 15:31:34	U	SMPL
CCV38 - 1	CCV38	18766	4.9340	-	7470A	PPB	02 Mar 2020 15:33:41	U	SMPL
CCB38 - 1	CCB38	-296	-0.0856	-	7470A	PPB	02 Mar 2020 15:35:48	U	SMPL
L1723-03 - 1	WC	-90	-0.0313	-	7470A	PPB	02 Mar 2020 15:37:56	U	SMPL
L1743-19 - 1	C-7	-37	-0.0174	-	7470A	PPB	02 Mar 2020 15:40:02	U	SMPL
L1743-20 - 1	C-8	-23	-0.0137	-	7470A	PPB	02 Mar 2020 15:42:08	U	SMPL
L1743-21 - 1	C-9	-27	-0.0147	-	7470A	PPB	02 Mar 2020 15:44:14	U	SMPL
L1743-22 - 1	C-3	-12	-0.0108	-	7470A	PPB	02 Mar 2020 15:46:21	U	SMPL
L1743-23 - 1	C-4	-1	-0.0079	-	7470A	PPB	02 Mar 2020 15:48:28	U	SMPL
L1743-24 - 1	C-5	-14	-0.0113	-	7470A	PPB	02 Mar 2020 15:50:35	U	SMPL
L1743-25 - 1	C-6	-8	-0.0097	-	7470A	PPB	02 Mar 2020 15:52:43	U	SMPL
L1743-26 - 1	C-1	-4	-0.0087	-	7470A	PPB	02 Mar 2020 15:54:50	U	SMPL
L1743-27 - 1	C-2	-14	-0.0113	-	7470A	PPB	02 Mar 2020 15:56:56	U	SMPL
CCV39 - 1	CCV39	18248	4.7976	-	7470A	PPB	02 Mar 2020 15:59:02	U	SMPL
CCB39 - 1	CCB39	-212	-0.0635	-	7470A	PPB	02 Mar 2020 16:01:08	U	SMPL
L1744-06 - 1	TP-3	-69	-0.0258	-	7470A	PPB	02 Mar 2020 16:03:15	U	SMPL
L1744-07 - 1	TP-2	-15	-0.0116	-	7470A	PPB	02 Mar 2020 16:05:21	U	SMPL
L1744-08 - 1	TP-1	-22	-0.0134	-	7470A	PPB	02 Mar 2020 16:07:28	U	SMPL
L1744-09 - 1	TP-4	-8	-0.0097	-	7470A	PPB	02 Mar 2020 16:09:34	U	SMPL
L1745-02 - 1	MP-2-COMP	-8	-0.0097	-	7470A	PPB	02 Mar 2020 16:11:41	U	SMPL
PB127210BL - 1	PBW	-11	-0.0105	-	7470A	PPB	02 Mar 2020 16:13:48	U	SMPL
PB127210BS - 1	LCSW	13568	3.5652	-	7470A	PPB	02 Mar 2020 16:15:55	U	SMPL
L1717-02 - 1	PLANK-MANHOLE	-183	-0.0558	-	7470A	PPB	02 Mar 2020 16:18:03	U	SMPL
L1724-02 - 1	A2828	-55	-0.0221	-	7470A	PPB	02 Mar 2020 16:20:10	U	SMPL
L1724-02DUP - 1	A2828DUP	-25	-0.0142	-	7470A	PPB	02 Mar 2020 16:22:17	U	SMPL
CCV40 - 1	CCV40	17928	4.7133	-	7470A	PPB	02 Mar 2020 16:24:23	U	SMPL
CCB40 - 1	CCB40	-255	-0.0748	-	7470A	PPB	02 Mar 2020 16:26:29	U	SMPL
L1724-02MS - 1	A2828MS	13271	3.4870	-	7470A	PPB	02 Mar 2020 16:31:32	U	SMPL
L1724-02MSD - 1	A2828MSD	13567	3.5649	-	7470A	PPB	02 Mar 2020 16:33:38	U	SMPL
PB127211BL - 1	PBW	-264	-0.0772	-	7470A	PPB	02 Mar 2020 16:35:45	U	SMPL
PB127211BS - 1	LCSW	12858	3.3782	-	7470A	PPB	02 Mar 2020 16:37:53	U	SMPL
L1724-01 - 1	A2828	-241	-0.0711	-	7470A	PPB	02 Mar 2020 16:39:59	U	SMPL
L1724-01DUP - 1	A2828DUP	87	0.0153	-	7470A	PPB	02 Mar 2020 16:42:07	U	SMPL
L1724-01MS - 1	A2828MS	13494	3.5457	-	7470A	PPB	02 Mar 2020 16:46:42	U	SMPL
L1724-01MSD - 1	A2828MSD	13595	3.5723	-	7470A	PPB	02 Mar 2020 16:48:49	U	SMPL
PB127209TB - 1	PB127165TB	-99	-0.0337	-	7470A	PPB	02 Mar 2020 16:50:57	U	SMPL
PB127210TB - 1	PB127167TB	4	-0.0066	-	7470A	PPB	02 Mar 2020 16:53:04	U	SMPL
CCV41 - 1	CCV41	18004	4.7333	-	7470A	PPB	02 Mar 2020 16:55:10	U	SMPL
CCB41 - 1	CCB41	-290	-0.0840	-	7470A	PPB	02 Mar 2020 16:57:16	U	SMPL
L1372-10LX5 - 1		-81	-0.0290	-	7470A	PPB	02 Mar 2020 16:59:23	U	SMPL
L1372-10A - 1		13161	3.4580	-	7470A	PPB	02 Mar 2020 17:01:29	U	SMPL
L1724-02LX5 - 1		-260	-0.0761	-	7470A	PPB	02 Mar 2020 17:03:35	U	SMPL
L1724-02A - 1		13429	3.5286	-	7470A	PPB	02 Mar 2020 17:05:43	U	SMPL
L1724-01LX5 - 1		-275	-0.0801	-	7470A	PPB	02 Mar 2020 17:07:49	U	SMPL
L1724-01A - 1		13270	3.4867	-	7470A	PPB	02 Mar 2020 17:09:57	U	SMPL
CCV42 - 1	CCV42	18508	4.8660	-	7470A	PPB	02 Mar 2020 17:12:05	U	SMPL
CCB42 - 1	CCB42	-330	-0.0945	-	7470A	PPB	02 Mar 2020 17:14:12	U	SMPL