



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108051

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/02/2020	17:40
2	CAL2	1	Water	NA	NA	20.2	7.00	03/02/2020	17:41
3	CAL3	1	Water	NA	NA	20.1	10.04	03/02/2020	17:42
4	ICV	1	Water	NA	NA	20.1	7.00	03/02/2020	17:43
5	CCV1	1	Water	NA	NA	20.2	2.01	03/02/2020	17:45
6	L1723-03	1	Solid	20.03	20	20.3	8.04	03/02/2020	17:46
7	L1743-19	1	Solid	20.04	20	20.0	8.31	03/02/2020	17:48
8	L1743-20	1	Solid	20.02	20	19.9	8.21	03/02/2020	17:49
9	L1743-21	1	Solid	20.03	20	19.9	7.82	03/02/2020	17:50
10	L1743-22	1	Solid	20.02	20	19.8	8.16	03/02/2020	18:00
11	L1743-23	1	Solid	20.04	20	20.3	8.45	03/02/2020	18:02
12	L1743-24	1	Solid	20.03	20	20.2	8.62	03/02/2020	18:05
13	L1743-25	1	Solid	20.02	20	20.2	8.65	03/02/2020	18:06
14	L1743-26	1	Solid	20.05	20	19.6	9.15	03/02/2020	18:09
15	L1743-27	1	Solid	20.02	20	20.2	7.59	03/02/2020	18:10
16	CCV2	1	Water	NA	NA	20.1	12.02	03/02/2020	18:11
17	L1757-06	1	Solid	20.03	20	20.4	8.62	03/02/2020	18:12
18	L1762-03	1	Solid	20.02	20	20.2	8.22	03/02/2020	18:14
19	L1762-06	1	Solid	20.04	20	20.1	8.03	03/02/2020	18:15
20	L1762-09	1	Solid	20.03	20	20.4	8.76	03/02/2020	18:16

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L1762-09DUP	1	Solid	20.04	20	20.5	8.77	03/02/2020	18:18
22	CCV3	1	Water	NA	NA	20.2	2.01	03/02/2020	18:20

VB108051

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1762 WorkList ID : 136633 Department : Wet-Chemistry Date : 03-02-2020 16:19:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2020	Solid	L1723-03	Corrosivity	Cool 4 deg C	DIST01	WC	WC	02/27/2020	9045D
02/28/2020	Solid	L1743-19	Corrosivity	Cool 4 deg C	DVIR01	C-7	C-7	02/28/2020	9045D
02/28/2020	Solid	L1743-20	Corrosivity	Cool 4 deg C	DVIR01	C-8	C-8	02/28/2020	9045D
02/28/2020	Solid	L1743-21	Corrosivity	Cool 4 deg C	DVIR01	C-9	C-9	02/28/2020	9045D
02/28/2020	Solid	L1743-22	Corrosivity	Cool 4 deg C	DVIR01	C-3	C-3	02/28/2020	9045D
02/28/2020	Solid	L1743-23	Corrosivity	Cool 4 deg C	DVIR01	C-4	C-4	02/28/2020	9045D
02/28/2020	Solid	L1743-24	Corrosivity	Cool 4 deg C	DVIR01	C-5	C-5	02/28/2020	9045D
02/28/2020	Solid	L1743-25	Corrosivity	Cool 4 deg C	DVIR01	C-6	C-6	02/28/2020	9045D
02/28/2020	Solid	L1743-26	Corrosivity	Cool 4 deg C	DVIR01	C-1	C-1	02/28/2020	9045D
02/28/2020	Solid	L1743-27	Corrosivity	Cool 4 deg C	DVIR01	C-2	C-2	02/28/2020	9045D
03/02/2020	Solid	L1757-06	Corrosivity	Cool 4 deg C	PSEG05	I61	EO-02-030220	03/02/2020	9045D
03/02/2020	Solid	L1762-03	Corrosivity	Cool 4 deg C	PSEG05	I62	TR-04-030220	03/02/2020	9045D
03/02/2020	Solid	L1762-06	Corrosivity	Cool 4 deg C	PSEG05	I62	TR-05-030220	03/02/2020	9045D
03/02/2020	Solid	L1762-09	Corrosivity	Cool 4 deg C	PSEG05	I62	TR-06-030220	03/02/2020	9045D

Date/Time 03.02.2020 16:19
 Received by: Jg
 Relinquished by: Jg

Date/Time 03.02.2020 18:25
 Received by: Jg
 Relinquished by: Jg