

UB108070

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

3/3/2020 14:54

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.719	0.0	0.078	
ICB1	0.893	0.0	0.003	
CCV1	244.031	0.0	0.198	
CCB1	0.526	0.0	0.003	
PB127225BL	0.406	0.0	0.003	
PB127225BS	94.650	0.0	0.078	
L1743-02	0.661	0.0	0.003	
L1743-02DUP	0.681	0.0	0.003	
L1743-04	0.453	0.0	0.003	
L1743-06	0.544	0.0	0.003	
L1743-08	0.465	0.0	0.003	
L1743-10	1.721	0.0	0.004	
L1743-12	0.844	0.0	0.003	
L1743-14	0.893	0.0	0.003	
CCV2	254.082	0.0	0.207	
CCB2	0.745	0.0	0.003	
L1743-16	2.118	0.0	0.004	
L1743-18	2.142	0.0	0.004	
CCV3	257.252	0.0	0.209	
CCB3	0.666	0.0	0.003	

N	20
Mean	47.875
SD	92.3970
CV%	193.00

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : FM

Instrument ID : Konelab

3/3/2020 11:47


Test Total CN

Accepted 3/3/2020 11:47

Factor 1244
Bias 0.002

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.8930	0.0000	
2	5.0PPBCN	0.007	5.4813	5.0000	
3	10PPBCN	0.011	10.5926	10.0000	
4	50PPBCN	0.042	48.9544	50.0000	
5	100PPBCN	0.081	97.5863	100.0000	
6	250PPBCN	0.205	251.8436	250.0000	
7	500PPBCN	0.404	499.6489	500.0000	

Aquakem v. 7.2AQ1

AD

Results from time period:

Tue Mar 03 14:22:21 2020

Tue Mar 03 14:52:11 2020

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	0.893	µg/l	3/3/2020 11:42:16	
5.0PPBCN	A	Total CN	P	5.4813	µg/l	3/3/2020 11:42:17	
10PPBCN	A	Total CN	P	10.5926	µg/l	3/3/2020 11:42:18	
50PPBCN	A	Total CN	P	48.9544	µg/l	3/3/2020 11:42:19	
100PPBCN	A	Total CN	P	97.5863	µg/l	3/3/2020 11:42:20	
250PPBCN	A	Total CN	P	251.8436	µg/l	3/3/2020 11:42:21	
500PPBCN	A	Total CN	P	499.6489	µg/l	3/3/2020 11:42:22	
ICV1	S	Total CN	P	93.7189	µg/l	3/3/2020 14:22:21	
ICB1	S	Total CN	P	0.8935	µg/l	3/3/2020 14:22:22	
CCV1	S	Total CN	P	244.0308	µg/l	3/3/2020 14:22:23	
CCB1	S	Total CN	P	0.526	µg/l	3/3/2020 14:22:24	
PB127225BL	S	Total CN	P	0.4064	µg/l	3/3/2020 14:22:25	
PB127225BS	S	Total CN	P	94.6503	µg/l	3/3/2020 14:22:26	
L1743-02	S	Total CN	P	0.6607	µg/l	3/3/2020 14:22:27	
L1743-02DUP	S	Total CN	P	0.6811	µg/l	3/3/2020 14:22:28	
L1743-04	S	Total CN	P	0.4532	µg/l	3/3/2020 14:22:29	
L1743-06	S	Total CN	P	0.544	µg/l	3/3/2020 14:22:30	
L1743-08	S	Total CN	P	0.4645	µg/l	3/3/2020 14:22:31	
L1743-10	S	Total CN	P	1.7213	µg/l	3/3/2020 14:29:00	
L1743-12	S	Total CN	P	0.8443	µg/l	3/3/2020 14:29:01	
L1743-14	S	Total CN	P	0.8932	µg/l	3/3/2020 14:29:02	
CCV2	S	Total CN	P	254.0822	µg/l	3/3/2020 14:52:06	
CCB2	S	Total CN	P	0.7455	µg/l	3/3/2020 14:52:07	
L1743-16	S	Total CN	P	2.1181	µg/l	3/3/2020 14:52:08	
L1743-18	S	Total CN	P	2.1423	µg/l	3/3/2020 14:52:09	
CCV3	S	Total CN	P	257.2519	µg/l	3/3/2020 14:52:10	
CCB3	S	Total CN	P	0.6659	µg/l	3/3/2020 14:52:11	