



Analytical Summary Report

Analysis Method: 9045D
Parameter: pH
Run Number: LB108074
BalanceID: WC SC-4

Analyst By : jignesh
Supervisor Review By : apatel
Slope : 98.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/03/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	7.00	03/03/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	03/03/2020	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	03/03/2020	16:15
5	CCV1	1	Water	NA	NA	20.1	2.01	03/03/2020	16:16
6	L1721-01	1	Solid	20.03	20	21.2	5.80	03/03/2020	16:19
7	L1721-01DUP	1	Solid	20.04	20	21.1	5.81	03/03/2020	16:20
8	CCV2	1	Water	NA	NA	20.1	12.02	03/03/2020	16:22

W108074

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1721

WorkList ID : 136669

Department : Wet-Chemistry

Date : 03-03-2020 15:50:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/27/2020	Solid	L1721-01	pH	Cool 4 deg C	TULL01	O31	STOCKPILE-GRIT	02/27/2020	9045D

Date/Time 03-03-2020 / 16:00
 Received by: JD
 Relinquished by: SA

Date/Time 03-03-2020 / 17:35
 Received by: SA
 Relinquished by: JD