



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108083

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/04/2020	10:35
2	L1717-02		1	8	100.00	>212.0	>212.0	03/04/2020	10:58
3	L1717-02DUP		1	8	100.00	>212.0	>212.0	03/04/2020	11:25
4	L1724-02		1	9	100.00	>212.0	>212.0	03/04/2020	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp3420

WorkList ID : 136692

Department : Wet-Chemistry

Date : 03-04-2020 10:01:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2020	Water	L1717-02	Flash Point	Cool 4 deg C	PSEG01	O32	PLANK-MANHOLE	02/27/2020	1010A
03/04/2020	Water	L1724-02	Flash Point	Cool 4 deg C	PSEG01	I42	A2828	02/28/2020	1010A

Date/Time 03/04/20 10:05
 Received by: KS
 Relinquished by: KS

Date/Time 03/04/20 11:55
 Received by: KS
 Relinquished by: KS