

LB 102084


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

51210



A= 0.0000e+000

B= 1.9667e-004

C= 4.2850e-002

Rho= 0.9997407

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	70	0.000	70				
0.2	0.200	0.217	0.017	887	0.0 %	887				
2.5	2.500	2.448	-0.052	12228	0.0 %	12228				
5.0	5.000	5.015	0.015	25284	0.0 %	25284				
7.5	7.500	7.349	-0.151	37149	0.0 %	37149				
10.0	10.000	10.114	0.114	51210	0.0 %	51210				

LB108084 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 04 Mar 2020 11:53:25				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date Type Type		
0.0 - 1	50	70	-	0.00007471B	PPB	04 Mar 2020 12:34:29	S	Std
0.2 - 1	50.2	887	-	0.20007471B	PPB	04 Mar 2020 12:36:46	S	Std
2.5 - 1	52.5	12228	-	2.50007471B	PPB	04 Mar 2020 12:39:03	S	Std
5.0 - 1	55	25284	-	5.00007471B	PPB	04 Mar 2020 12:41:20	S	Std
7.5 - 1	57.5	37149	-	7.50007471B	PPB	04 Mar 2020 12:43:36	S	Std
10.0 - 1	510	51210	-	10.00007471B	PPB	04 Mar 2020 12:45:54	S	Std
ICV44 - 1	ICV44	19733	3.9237	-7471B	PPB	04 Mar 2020 12:48:43	U	SMPL
ICB44 - 1	ICB44	-307	-0.0175	-7471B	PPB	04 Mar 2020 12:50:59	U	SMPL
CCV52 - 1	CCV52	25374	5.0331	-7471B	PPB	04 Mar 2020 12:53:16	U	SMPL
CCB52 - 1	CCB52	-330	-0.0221	-7471B	PPB	04 Mar 2020 12:55:30	U	SMPL
CRA - 1	CRA	830	0.2061	-7471B	PPB	04 Mar 2020 12:57:48	U	SMPL
HighStd - 1	HighStd	50526	9.9797	-7471B	PPB	04 Mar 2020 13:00:03	U	SMPL
ChkStd - 1	ChkStd	36790	7.2782	-7471B	PPB	04 Mar 2020 13:02:18	U	SMPL
PB127252BL - 1	PBS	-364	-0.0287	-7471B	PPB	04 Mar 2020 13:04:36	U	SMPL
PB127252BS - 1	LCSS	18934	3.7665	-7471B	PPB	04 Mar 2020 13:06:54	U	SMPL
L1721-01 - 1	STOCKPILE-GRIT	14526	2.8996	-7471B	PPB	04 Mar 2020 13:09:10	U	SMPL
L1748-01 - 1	AU-06-030320	3208	0.6738	-7471B	PPB	04 Mar 2020 13:11:27	U	SMPL
L1759-01 - 1	OR-02-030320	1362	0.3107	-7471B	PPB	04 Mar 2020 13:13:46	U	SMPL
L1771-01 - 1	PIER-122-CORE-1-3-2-20	7324	1.4832	-7471B	PPB	04 Mar 2020 13:16:04	U	SMPL
L1771-02 - 1	PIER-122-CORE-2-3-2-20	8249	1.6652	-7471B	PPB	04 Mar 2020 13:18:19	U	SMPL
CCV53 - 1	CCV53	26316	5.2183	-7471B	PPB	04 Mar 2020 13:20:35	U	SMPL
CCB53 - 1	CCB53	-122	0.0189	-7471B	PPB	04 Mar 2020 13:22:50	U	SMPL
L1774-02 - 1	VNJ-223	2586	0.5514	-7471B	PPB	04 Mar 2020 13:25:06	U	SMPL
L1774-02DUP - 1	VNJ-223DUP	2140	0.4637	-7471B	PPB	04 Mar 2020 13:27:21	U	SMPL
L1774-02MS - 1	VNJ-223MS	20805	4.1345	-7471B	PPB	04 Mar 2020 13:29:37	U	SMPL
L1774-02MSD - 1	VNJ-223MSD	20951	4.1632	-7471B	PPB	04 Mar 2020 13:31:52	U	SMPL
L1774-04 - 1	VNJ-236	2020	0.4401	-7471B	PPB	04 Mar 2020 13:34:09	U	SMPL
L1775-01 - 1	CR-1	2759	0.5855	-7471B	PPB	04 Mar 2020 13:36:27	U	SMPL
L1775-07 - 1	CR-2	1561	0.3498	-7471B	PPB	04 Mar 2020 13:38:43	U	SMPL
L1775-13 - 1	CR-3	1655	0.3683	-7471B	PPB	04 Mar 2020 13:41:00	U	SMPL
L1774-02LX5 - 1		295	0.1009	-7471B	PPB	04 Mar 2020 13:43:17	U	SMPL
L1774-02A - 1		23757	4.7151	-7471B	PPB	04 Mar 2020 13:45:32	U	SMPL
CCV54 - 1	CCV54	24583	4.8775	-7471B	PPB	04 Mar 2020 13:47:48	U	SMPL
CCB54 - 1	CCB54	-145	0.0143	-7471B	PPB	04 Mar 2020 13:50:04	U	SMPL