



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108098

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/04/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	03/04/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	03/04/2020	16:03
4	ICV	1	Water	NA	NA	20.2	7.00	03/04/2020	16:05
5	CCV1	1	Water	NA	NA	20.1	2.01	03/04/2020	16:06
6	L1777-01	1	Solid	20.03	20	20.1	6.41	03/04/2020	16:10
7	L1777-01DUP	1	Solid	20.04	20	21.2	6.42	03/04/2020	16:11
8	L1777-02	1	Solid	20.02	20	20.3	5.48	03/04/2020	16:15
9	CCV2	1	Water	NA	NA	20.1	12.02	03/04/2020	16:16

78 108098

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1777 WorkList ID : 136701 Department : Wet-Chemistry Date : 03-04-2020 15:40:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/04/2020	Solid	L1777-01	pH	Cool 4 deg C	PSEG01	I41	40681	03/04/2020	9045D
03/04/2020	Solid	L1777-02	pH	Cool 4 deg C	PSEG01	I41	40683	03/04/2020	9045D

Date/Time 03-04-2020 15:45:00
Received by: JA
Relinquished by: JA

Date/Time 03-04-2020 16:15:00
Received by: JA
Relinquished by: JA