


LB108110

7471B

INSTRUMENT ID: CY4



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.037	-0.037	247	0.000	247				
0.2	0.200	0.140	-0.060	1230	0.0 %	1230				
2.5	2.500	2.540	0.040	14574	0.0 %	14574				
5.0	5.000	5.074	0.074	28662	0.0 %	28662				
7.5	7.500	7.639	0.139	42925	0.0 %	42925				
10.0	10.000	9.845	-0.155	55190	0.0 %	55190				

LB108110 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Mar 2020 10:38:00

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	247	-	0.0000	7471B	PPB	05 Mar 2020 10:40:42	S	Std
0.2 - 1	50.2	1230	-	0.2000	7471B	PPB	05 Mar 2020 10:42:59	S	Std
2.5 - 1	52.5	14574	-	2.5000	7471B	PPB	05 Mar 2020 10:45:16	S	Std
5.0 - 1	55	28662	-	5.0000	7471B	PPB	05 Mar 2020 10:47:33	S	Std
7.5 - 1	57.5	42925	-	7.5000	7471B	PPB	05 Mar 2020 10:49:50	S	Std
10.0 - 1	510	55190	-	10.0000	7471B	PPB	05 Mar 2020 11:00:14	S	Std
ICV46 - 1	ICV46	20520	3.6092	-	7471B	PPB	05 Mar 2020 11:05:54	U	SMPL
ICB46 - 1	ICB46	-17	-0.0844	-	7471B	PPB	05 Mar 2020 11:08:09	U	SMPL
CCV58 - 1	CCV58	27199	4.8105	-	7471B	PPB	05 Mar 2020 11:10:25	U	SMPL
CCB58 - 1	CCB58	-62	-0.0925	-	7471B	PPB	05 Mar 2020 11:12:40	U	SMPL
CRA - 1	CRA	1294	0.1514	-	7471B	PPB	05 Mar 2020 11:17:28	U	SMPL
HighStd - 1	HighStd	55537	9.9072	-	7471B	PPB	05 Mar 2020 11:19:42	U	SMPL
ChkStd - 1	ChkStd	39602	7.0412	-	7471B	PPB	05 Mar 2020 11:21:57	U	SMPL
PB127273BL - 1	PBS	692	0.0431	-	7471B	PPB	05 Mar 2020 11:26:54	U	SMPL
PB127273BS - 1	LCSS	20187	3.5494	-	7471B	PPB	05 Mar 2020 11:29:10	U	SMPL
L1777-03 - 1	40681-40683	939	0.0875	-	7471B	PPB	05 Mar 2020 11:31:26	U	SMPL
L1778-01 - 1	S1	50562	9.0124	-	7471B	PPB	05 Mar 2020 11:33:44	U	SMPL
L1778-01DUP - 1	S1DUP	49203	8.7680	-	7471B	PPB	05 Mar 2020 11:36:01	U	SMPL
L1778-01MS - 1	S1MS	77018	13.7706	-	7471B	PPB	05 Mar 2020 11:40:57	U	SMPL
L1778-01MSD - 1	S1MSD	77363	13.8327	-	7471B	PPB	05 Mar 2020 11:43:14	U	SMPL
CCV59 - 1	CCV59	27204	4.8114	-	7471B	PPB	05 Mar 2020 11:45:31	U	SMPL
CCB59 - 1	CCB59	1	-0.0812	-	7471B	PPB	05 Mar 2020 11:47:48	U	SMPL
L1778-03 - 1	S2	38157	6.7813	-	7471B	PPB	05 Mar 2020 11:50:04	U	SMPL
L1778-05 - 1	DN1	20163	3.5450	-	7471B	PPB	05 Mar 2020 11:52:19	U	SMPL
L1778-07 - 1	DN2	3224	0.4985	-	7471B	PPB	05 Mar 2020 11:54:36	U	SMPL
L1778-09 - 1	S3	7756	1.3136	-	7471B	PPB	05 Mar 2020 11:56:52	U	SMPL
L1778-11 - 1	S4	11664	2.0165	-	7471B	PPB	05 Mar 2020 11:59:08	U	SMPL
L1778-13 - 1	S5	6064	1.0093	-	7471B	PPB	05 Mar 2020 12:01:24	U	SMPL
L1778-15 - 1	S7	8461	1.4404	-	7471B	PPB	05 Mar 2020 12:03:40	U	SMPL
L1778-17 - 1	S6	20611	3.6256	-	7471B	PPB	05 Mar 2020 12:05:57	U	SMPL
L1778-19 - 1	S11	7543	1.2753	-	7471B	PPB	05 Mar 2020 12:08:13	U	SMPL
L1778-21 - 1	S10	4676	0.7596	-	7471B	PPB	05 Mar 2020 12:10:29	U	SMPL
CCV60 - 1	CCV60	27621	4.8864	-	7471B	PPB	05 Mar 2020 12:12:44	U	SMPL
CCB60 - 1	CCB60	377	-0.0135	-	7471B	PPB	05 Mar 2020 12:14:59	U	SMPL
L1778-23 - 1	S9	3524	0.5525	-	7471B	PPB	05 Mar 2020 12:17:16	U	SMPL
L1778-25 - 1	S8	2301	0.3325	-	7471B	PPB	05 Mar 2020 12:19:31	U	SMPL
L1783-01 - 1	GP-10-030220	2082	0.2931	-	7471B	PPB	05 Mar 2020 12:21:47	U	SMPL
L1783-02 - 1	GP-11-030220	1309	0.1541	-	7471B	PPB	05 Mar 2020 12:24:02	U	SMPL
L1797-01 - 1	6-FEET	13110	2.2765	-	7471B	PPB	05 Mar 2020 12:26:18	U	SMPL
L1797-02 - 1	30-FEET	30632	5.4279	-	7471B	PPB	05 Mar 2020 12:28:34	U	SMPL
PB127277BL - 1	PBS	393	-0.0107	-	7471B	PPB	05 Mar 2020 12:30:51	U	SMPL
PB127277BS - 1	LCSS	19900	3.4977	-	7471B	PPB	05 Mar 2020 12:33:09	U	SMPL
L1761-01 - 1	SU-04-030420	2926	0.4449	-	7471B	PPB	05 Mar 2020 12:40:34	U	SMPL
L1761-01DUP - 1	SU-04-030420DUP	2872	0.4352	-	7471B	PPB	05 Mar 2020 12:42:50	U	SMPL
CCV61 - 1	CCV61	28229	4.9957	-	7471B	PPB	05 Mar 2020 12:45:05	U	SMPL
CCB61 - 1	CCB61	501	0.0088	-	7471B	PPB	05 Mar 2020 12:47:20	U	SMPL
L1761-01MS - 1	SU-04-030420MS	23286	4.1067	-	7471B	PPB	05 Mar 2020 12:49:37	U	SMPL
L1761-01MSD - 1	SU-04-030420MSD	23180	4.0877	-	7471B	PPB	05 Mar 2020 12:51:52	U	SMPL
L1792-01 - 1	TP-3	1423	0.1746	-	7471B	PPB	05 Mar 2020 12:54:08	U	SMPL
L1792-03 - 1	TP-4	2987	0.4559	-	7471B	PPB	05 Mar 2020 12:56:24	U	SMPL
L1792-05 - 1	TP-5	4323	0.6962	-	7471B	PPB	05 Mar 2020 12:58:40	U	SMPL
L1792-07 - 1	TP-2	2498	0.3679	-	7471B	PPB	05 Mar 2020 13:00:56	U	SMPL
L1792-09 - 1	TP-1	2786	0.4197	-	7471B	PPB	05 Mar 2020 13:03:13	U	SMPL
CCV62 - 1	CCV62	29631	5.2479	-	7471B	PPB	05 Mar 2020 13:05:30	U	SMPL
CCB62 - 1	CCB62	514	0.0111	-	7471B	PPB	05 Mar 2020 13:07:47	U	SMPL
L1778-01LX5 - 1		9722	1.6672	-	7471B	PPB	05 Mar 2020 13:10:03	U	SMPL
L1778-01A - 1		74801	13.3719	-	7471B	PPB	05 Mar 2020 13:12:19	U	SMPL
L1761-01LX5 - 1		515	0.0113	-	7471B	PPB	05 Mar 2020 13:14:34	U	SMPL
L1761-01A - 1		27187	4.8083	-	7471B	PPB	05 Mar 2020 13:20:01	U	SMPL
L1778-01DLX2 - 1		24509	4.3267	-	7471B	PPB	05 Mar 2020 13:22:17	U	SMPL
L1778-01DUPDLX2 - 1		24364	4.3006	-	7471B	PPB	05 Mar 2020 13:24:33	U	SMPL
L1778-01MSDLX2 - 1		40076	7.1265	-	7471B	PPB	05 Mar 2020 13:26:49	U	SMPL
L1778-01MSDDLX2 - 1		38969	6.9274	-	7471B	PPB	05 Mar 2020 13:29:05	U	SMPL
L1778-01LDLX10 - 1		4905	0.8008	-	7471B	PPB	05 Mar 2020 13:31:22	U	SMPL
L1778-01ADLX2 - 1		37927	6.7400	-	7471B	PPB	05 Mar 2020 13:33:40	U	SMPL
CCV63 - 1	CCV63	28417	5.0295	-	7471B	PPB	05 Mar 2020 13:35:56	U	SMPL
CCB63 - 1	CCB63	170	-0.0508	-	7471B	PPB	05 Mar 2020 13:38:14	U	SMPL