



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108112

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/05/2020	11:00
2	CAL2	1	Water	NA	NA	20.1	7.00	03/05/2020	11:01
3	CAL3	1	Water	NA	NA	20.1	10.03	03/05/2020	11:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/05/2020	11:05
5	CCV1	1	Water	NA	NA	20.2	2.01	03/05/2020	11:06
6	L1807-01	1	Water	NA	NA	18.1	5.77	03/05/2020	11:10
7	L1807-01DUP	1	Water	NA	NA	18.2	5.78	03/05/2020	11:11
8	L1808-01	1	Water	NA	NA	19.3	4.92	03/05/2020	11:15
9	L1808-02	1	Water	NA	NA	19.9	4.75	03/05/2020	11:16
10	L1808-03	1	Water	NA	NA	19.8	4.76	03/05/2020	11:19
11	CCV2	1	Water	NA	NA	20.1	12.02	03/05/2020	11:20

WORKLIST(Hardcopy Internal Chain)

108112

WorkList Name : PH L1807 WorkList ID : 136744 Department : Wet-Chemistry Date : 03-05-2020 10:40:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/02/2020	Water	L1807-01-✓	pH	Cool 4 deg C	STER14	O31	JS-10000-SYSTEM	03/02/2020	SM 4500-H B
03/04/2020	Water	L1808-01-✓	pH	Cool 4 deg C	STER13	O31	CELL-A-POOL-WATER	03/04/2020	SM 4500-H B
03/04/2020	Water	L1808-02-✓	pH	Cool 4 deg C	STER13	O31	CELL-B-POOL-WATER	03/04/2020	SM 4500-H B
03/04/2020	Water	L1808-03-✓	pH	Cool 4 deg C	STER13	O31	CELL-C-POOL-WATER	03/04/2020	SM 4500-H B

Date/Time 03-05-2020 10:50
 Received by: JG
 Relinquished by: JG

Date/Time 03-05-2020 12:10
 Received by: JG
 Relinquished by: JG