



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108116

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/05/2020	13:15
2	CAL2	1	Water	NA	NA	20.1	6.99	03/05/2020	13:16
3	CAL3	1	Water	NA	NA	20.1	10.03	03/05/2020	13:18
4	ICV	1	Water	NA	NA	20.2	7.00	03/05/2020	13:20
5	CCV1	1	Water	NA	NA	20.1	2.01	03/05/2020	13:22
6	L1813-01	1	Solid	20.03	20	21.0	6.58	03/05/2020	13:25
7	L1813-01DUP	1	Solid	20.04	20	21.1	6.59	03/05/2020	13:26
8	CCV2	1	Water	NA	NA	20.1	12.02	03/05/2020	13:30

WORKLIST(Hardcopy Internal Chain)

108116

WorkList Name : PH L1813 WorkList ID : 136755 Department : Wet-Chemistry Date : 03-05-2020 13:01:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2020	Solid	L1813-01-1	pH	Cool 4 deg C	HOME01	O41	50-FAIRVIEW-AVE-SECAUCL	03/05/2020	9045D

Date/Time 03-05-2020 13:10
 Received by: JG
 Relinquished by: JG

Date/Time 03-05-2020 14:30
 Received by: JG
 Relinquished by: JG