



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108121

Reagent/Standard	Lot/Log #
Residual chlorine std, and CCV , 0.4PPM	WP82731
Residual chlorine ICV-LCS, 0.4PPM	WP82712
Chlorine Calibration std, 0.1ppm	WP82714
Chlorine Calibration std, 0.2ppm	WP82715
Chlorine Calibration std, 0.8ppm	WP82717
Chlorine Calibration std, 0.0ppm	WP82719
Chlorine Calibration std, 1.6ppm	WP82718
Total Chlorine Powder Pillows	W2249

Intercept: 0.0131

Slope: 1.0294

Regression: 0.999812

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	03/05/2020	13:13
2	CAL2	0.1	1	0.000	0.110	0.110	03/05/2020	13:15
3	CAL3	0.2	1	0.000	0.220	0.220	03/05/2020	13:16
4	CAL4	0.4	1	0.000	0.440	0.440	03/05/2020	13:18
5	CAL5	0.8	1	0.000	0.850	0.850	03/05/2020	13:20
6	CAL6	1.6	1	0.000	1.650	1.650	03/05/2020	13:21



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108121

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4300	0.4300	0.4050	03/05/2020	13:23
2	ICB				1	0.0000	0.0000	0.0000	-0.0130	03/05/2020	13:24
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.3950	03/05/2020	13:25
4	CCB1				1	0.0000	0.0000	0.0000	-0.0130	03/05/2020	13:27
5	LB108121BL	50	50		1	0.0000	0.0000	0.0000	-0.0130	03/05/2020	13:28
6	LB108121BS	50	50	0.4	1	0.0000	0.4300	0.4300	0.4050	03/05/2020	13:30
7	L1801-19	50	50		1	0.0000	0.7500	0.7500	0.7160	03/05/2020	13:32
8	L1801-19DUP	50	50		1	0.0000	0.7300	0.7300	0.6960	03/05/2020	13:35
9	CCV2			0.4	1	0.0000	0.4400	0.4400	0.4150	03/05/2020	13:36
10	CCB2				1	0.0000	0.0000	0.0000	-0.0130	03/05/2020	13:38