

LB108128


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43109



A= 0.0000e+000

B= 2.3333e-004

C= 1.5365e-002

Rho= 0.9998972

Accept=Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.001	-0.001	-70	0.000	-70					1
0.05	0.050	0.050	0.000	149	0.0 %	149					0
0.20	0.200	0.206	0.006	816	0.0 %	816					3
2.50	2.500	2.563	0.063	10919	0.0 %	10919					3
5.00	5.000	4.941	-0.059	21112	0.0 %	21112					1
7.50	7.500	7.417	-0.083	31721	0.0 %	31721					1
10.0	10.000	10.074	0.074	43109	0.0 %	43109					1

LB108128

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0	S0.0	-70	-	0	ISM02.4	PPB	3/5/2020 17:06	Std
0.05	S0.05	149	-	0.05	ISM02.4	PPB	3/5/2020 17:08	Std
0.2	S0.2	816	-	0.2	ISM02.4	PPB	3/5/2020 17:13	Std
2.5	S2.5	10919	-	2.5	ISM02.4	PPB	3/5/2020 17:15	Std
5	S5.0	21112	-	5	ISM02.4	PPB	3/5/2020 17:18	Std
7.5	S7.5	31721	-	7.5	ISM02.4	PPB	3/5/2020 17:20	Std
10	S10.0	43109	-	10	ISM02.4	PPB	3/5/2020 17:22	Std
ICV	ICV	16171	3.7885 -		ISM02.4	PPB	3/5/2020 17:25	SMPL
ICB	ICB	-3	0.0147 -		ISM02.4	PPB	3/5/2020 17:30	SMPL
CCV067	CCV067	21194	4.9605 -		ISM02.4	PPB	3/5/2020 17:32	SMPL
CCB067	CCB067	-4	0.0144 -		ISM02.4	PPB	3/5/2020 17:37	SMPL
PB127304BL	PBW304	18	0.0196 -		ISM02.4	PPB	3/5/2020 17:39	SMPL
L1789-01	METC43	404	0.1096 -		ISM02.4	PPB	3/5/2020 17:41	SMPL
L1789-02	METC44	-97	-0.0073 -		ISM02.4	PPB	3/5/2020 17:44	SMPL
L1789-03	METC45	-12	0.0126 -		ISM02.4	PPB	3/5/2020 17:46	SMPL
L1789-04	METC45D	18	0.0196 -		ISM02.4	PPB	3/5/2020 17:48	SMPL
L1789-05	METC45S	3808	0.9039 -		ISM02.4	PPB	3/5/2020 17:50	SMPL
CCV068	CCV068	20327	4.7582 -		ISM02.4	PPB	3/5/2020 17:53	SMPL
CCB068	CCB068	2	0.0158 -		ISM02.4	PPB	3/5/2020 17:57	SMPL