

Analysis Method: 120.1ANALYST: AlexanderParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB108130Meter ID: WC Conductance meter-1Thermometer ID: PH meter- temp probe-1

Reagent/Standard	Lot/Log #
conductivity standard	WP81395

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB108130BL		1	22.1	0.80	0.85	03/05/2020	15:20
2	LB108130BS	1413	1	21.7	1312.00	1400.26	03/05/2020	15:22
3	L1801-08		1	21.8	778.00	828.65	03/05/2020	15:24

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--