



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108136

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/05/2020	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/05/2020	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/05/2020	16:02
4	ICV	1	Water	NA	NA	20.2	7.01	03/05/2020	16:03
5	CCV1	1	Water	NA	NA	20.1	2.01	03/05/2020	16:05
6	L1753-08	1	Solid	20.03	20	21.6	8.40	03/05/2020	16:06
7	L1761-05	1	Solid	20.04	20	21.1	8.21	03/05/2020	16:09
8	L1762-12	1	Solid	20.02	20	21.5	8.19	03/05/2020	16:10
9	L1762-12DUP	1	Solid	20.03	20	21.6	8.20	03/05/2020	16:11
10	CCV2	1	Water	NA	NA	20.2	12.02	03/05/2020	16:15

WORKLIST(Hardcopy Internal Chain)

VB 108136

WorkList Name : CORROSIVITY L1761

WorkList ID : 136750

Department : Wet-Chemistry

Date : 03-05-2020 11:44:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2020	Solid	L1753-08	Corrosivity	Cool 4 deg C	PSEG05	I42	CL-01-030520	03/05/2020	9045D
03/04/2020	Solid	L1761-05	Corrosivity	Cool 4 deg C	PSEG05	I41	SU-04-030420	03/04/2020	9045D
03/05/2020	Solid	L1762-12	Corrosivity	Cool 4 deg C	PSEG05		TR-08-030520	03/05/2020	9045D

Date/Time 03-05-2020 1145
 Received by: Jg
 Relinquished by: Jg

Date/Time 03-05-2020 1840
 Received by: Jg
 Relinquished by: Jg