



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108137

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/05/2020	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	03/05/2020	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/05/2020	16:02
4	ICV	1	Water	NA	NA	20.2	7.01	03/05/2020	16:03
5	CCV1	1	Water	NA	NA	20.2	2.01	03/05/2020	17:15
6	L1803-02	1	Water	NA	NA	13.9	7.30	03/05/2020	17:16
7	L1803-02DUP	1	Water	NA	NA	13.8	7.31	03/05/2020	17:17
8	CCV2	1	Water	NA	NA	20.1	12.02	03/05/2020	17:18

WORKLIST(Hardcopy Internal Chain)

78108137

WorkList Name : CORROSIONITY L1803 WorkList ID : 136763 Department : Wet-Chemistry Date : 03-05-2020 17:11:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2020	Water	L1803-02-17	Corrosivity	Cool 4 deg C	PSEG01	L41	N38129	03/05/2020	9040C

Date/Time 03-05-2020 17:15
 Received by: Jø
 Relinquished by: Jø

Date/Time 03-05-2020 18:15
 Received by: Jø
 Relinquished by: Jø